

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079644.D
 Acq On : 23 Oct 2023 13:23
 Operator : JC\MD
 Sample : PB156545ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156545ZHE#04

Quant Time: Oct 23 23:49:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	187688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	391354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	411886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	106174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	167362	54.995	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.000%
35) Dibromofluoromethane	8.171	113	141476	50.304	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.600%
50) Toluene-d8	10.571	98	536677	52.739	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.480%
62) 4-Bromofluorobenzene	12.853	95	151487	46.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.780%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	17300	14.333	ug/l #	84
20) Methylene Chloride	5.271	84	17506	6.094	ug/l	88
37) Ethyl Acetate	7.577	43	4750	1.131	ug/l #	80
43) Isopropyl Acetate	8.694	43	182564	24.810	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079644.D
Acq On : 23 Oct 2023 13:23
Operator : JC\MD
Sample : PB156545ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156545ZHE#04

Quant Time: Oct 23 23:49:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

