

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079645.D
 Acq On : 23 Oct 2023 13:47
 Operator : JC\MD
 Sample : PB156545ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156545ZHE#05

Quant Time: Oct 23 23:50:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	180127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	376428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	386605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	166275	56.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.860%
35) Dibromofluoromethane	8.171	113	136650	50.515	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.020%
50) Toluene-d8	10.571	98	509798	52.084	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.160%
62) 4-Bromofluorobenzene	12.853	95	139667	44.463	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.920%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	17178	14.829	ug/l #	81
20) Methylene Chloride	5.289	84	17586	6.408	ug/l #	92
43) Isopropyl Acetate	8.694	43	187255	26.457	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079645.D
Acq On : 23 Oct 2023 13:47
Operator : JC\MD
Sample : PB156545ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156545ZHE#05

Quant Time: Oct 23 23:50:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

