

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079647.D
 Acq On : 23 Oct 2023 14:35
 Operator : JC\MD
 Sample : 04960-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VAUX-SLUDGE

Quant Time: Oct 23 23:50:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	245245	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	492086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	510045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	153161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	214153	53.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.720%	
35) Dibromofluoromethane	8.171	113	190955	53.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.000%	
50) Toluene-d8	10.571	98	614079	47.992	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.980%	
62) 4-Bromofluorobenzene	12.853	95	204542	49.812	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.620%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	20015	12.690	ug/l	91
20) Methylene Chloride	5.283	84	13211	3.262	ug/l #	75
43) Isopropyl Acetate	8.694	43	128480	13.886	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079647.D
Acq On : 23 Oct 2023 14:35
Operator : JC\MD
Sample : 04960-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VAUX-SLUDGE

Quant Time: Oct 23 23:50:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

