

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079648.D
 Acq On : 23 Oct 2023 14:59
 Operator : JC\MD
 Sample : 04976-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-426-COMP-01

Quant Time: Oct 23 23:51:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	231049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	478890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	500522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	152283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	207149	55.295	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.580%
35) Dibromofluoromethane	8.171	113	185284	53.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.680%
50) Toluene-d8	10.571	98	604815	48.571	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.140%
62) 4-Bromofluorobenzene	12.853	95	200437	50.157	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.320%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	19364	13.032	ug/l	88
20) Methylene Chloride	5.277	84	16170	4.421	ug/l	90
43) Isopropyl Acetate	8.694	43	186447	20.707	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079648.D
Acq On : 23 Oct 2023 14:59
Operator : JC\MD
Sample : 04976-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-426-COMP-01

Quant Time: Oct 23 23:51:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

