

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
 Data File : VN079651.D
 Acq On : 23 Oct 2023 16:12
 Operator : JC\MD
 Sample : 04978-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-406-COMP-03

Quant Time: Oct 23 23:52:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	227068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	475990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	496698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	204877	55.647	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.300%
35) Dibromofluoromethane	8.171	113	183753	53.719	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.440%
50) Toluene-d8	10.570	98	588555	47.553	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.100%
62) 4-Bromofluorobenzene	12.853	95	197521	49.729	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.460%
Target Compounds						
						Qvalue
16) Acetone	4.441	43	18246	12.495	ug/l	89
20) Methylene Chloride	5.277	84	14692	4.041	ug/l #	76
43) Isopropyl Acetate	8.694	43	209653	23.426	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102323\
Data File : VN079651.D
Acq On : 23 Oct 2023 16:12
Operator : JC\MD
Sample : 04978-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-406-COMP-03

Quant Time: Oct 23 23:52:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

