

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102418\
 Data File : VN051988.D
 Acq On : 24 Oct 2018 15:20
 Operator : MD\SY
 Sample : PB114120ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114120ZHE#04

Quant Time: Oct 25 02:10:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1084913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1577853	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1460436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	642371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	568253	58.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.62%	
35) Dibromofluoromethane	7.59	113	589827	58.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.16%	
50) Toluene-d8	10.09	98	2161242	59.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.80%	
62) 4-Bromofluorobenzene	12.40	95	665857	53.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.48%	
Target Compounds						
16) Acetone	3.82	43	7563	4.751	ug/l	93
20) Methylene Chloride	4.56	84	14166	1.517	ug/l	94
43) Isopropyl Acetate	8.17	43	280955	27.717	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102418\
Data File : VN051988.D
Acq On : 24 Oct 2018 15:20
Operator : MD\SY
Sample : PB114120ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114120ZHE#04

Quant Time: Oct 25 02:10:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

