

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102418\
 Data File : VN051989.D
 Acq On : 24 Oct 2018 15:47
 Operator : MD\SY
 Sample : PB114120ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114120ZHE#05

Quant Time: Oct 25 02:19:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1098755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1623265	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1501981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	660789	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	579668	59.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.48%	
35) Dibromofluoromethane	7.59	113	600091	57.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.86%	
50) Toluene-d8	10.09	98	2203764	58.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.74%	
62) 4-Bromofluorobenzene	12.40	95	684234	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
Target Compounds						
16) Acetone	3.82	43	8525	5.288	ug/l	99
20) Methylene Chloride	4.55	84	14092	1.490	ug/l	95
43) Isopropyl Acetate	8.17	43	273375	26.215	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102418\
Data File : VN051989.D
Acq On : 24 Oct 2018 15:47
Operator : MD\SY
Sample : PB114120ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114120ZHE#05

Quant Time: Oct 25 02:19:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

