

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102418\
 Data File : VN051994.D
 Acq On : 24 Oct 2018 18:00
 Operator : MD\SY
 Sample : PB114120ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114120ZHE#10

Quant Time: Oct 25 02:28:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1069217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1559967	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1474906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	679083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	552765	58.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.10%	
35) Dibromofluoromethane	7.59	113	577320	58.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.00%	
50) Toluene-d8	10.09	98	2154531	59.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.80%	
62) 4-Bromofluorobenzene	12.40	95	682887	55.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.46%	
Target Compounds						
16) Acetone	3.82	43	7114	4.534	ug/l #	79
43) Isopropyl Acetate	8.16	43	263296	26.273	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102418\
Data File : VN051994.D
Acq On : 24 Oct 2018 18:00
Operator : MD\SY
Sample : PB114120ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114120ZHE#10

Quant Time: Oct 25 02:28:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

