

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102418\
 Data File : VN051996.D
 Acq On : 24 Oct 2018 18:54
 Operator : MD\SY
 Sample : PB114120ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114120ZHE#06

Quant Time: Oct 25 02:32:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1148348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1696680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1603142	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	728736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	597951	58.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.94%	
35) Dibromofluoromethane	7.59	113	626546	57.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.74%	
50) Toluene-d8	10.09	98	2338512	59.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.54%	
62) 4-Bromofluorobenzene	12.40	95	738355	54.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.82%	
Target Compounds						
16) Acetone	3.82	43	6554	3.890	ug/l	91
43) Isopropyl Acetate	8.17	43	267812	24.570	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102418\
Data File : VN051996.D
Acq On : 24 Oct 2018 18:54
Operator : MD\SY
Sample : PB114120ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114120ZHE#06

Quant Time: Oct 25 02:32:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

