

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102419\
 Data File : VN058919.D
 Acq On : 23 Oct 2019 11:22
 Operator : JC/SP
 Sample : K5494-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-102119

Manual Integrations
 APPROVED

MMDadoda
 10/25/2019 2:44:32 PM

Quant Time: Oct 24 01:29:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	324479	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	517191	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	456906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	165135	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	242126	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	7.57	113	173993	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
50) Toluene-d8	10.09	98	637982	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	12.40	95	200832	41.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.86%	
Target Compounds						
16) Acetone	3.80	43	14410	6.128	ug/l	95
20) Methylene Chloride	4.52	84	5293m	1.318	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102419\
Data File : VN058919.D
Acq On : 23 Oct 2019 11:22
Operator : JC/SP
Sample : K5494-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-102119

Manual Integrations
APPROVED

MMDadoda
10/25/2019 2:44:32 PM

Quant Time: Oct 24 01:29:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

