

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102419\
 Data File : VN058920.D
 Acq On : 23 Oct 2019 11:44
 Operator : JC/SP
 Sample : PB124186TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124186TB

Quant Time: Oct 24 01:31:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	310534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	496822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	464026	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	171513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	231473	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
35) Dibromofluoromethane	7.57	113	164442	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
50) Toluene-d8	10.08	98	642059	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	12.40	95	204547	43.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.86%	
Target Compounds						
16) Acetone	3.80	43	16120	7.163	ug/l	91
20) Methylene Chloride	4.53	84	35832	9.324	ug/l	96
43) Isopropyl Acetate	8.15	43	88752	9.985	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102419\
Data File : VN058920.D
Acq On : 23 Oct 2019 11:44
Operator : JC/SP
Sample : PB124186TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124186TB

Quant Time: Oct 24 01:31:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

