

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102419\
 Data File : VN058927.D
 Acq On : 23 Oct 2019 14:18
 Operator : JC/SP
 Sample : K5139-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-03-102119

Quant Time: Oct 24 01:59:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	332535	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	527592	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	501707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	189515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	247132	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
35) Dibromofluoromethane	7.57	113	176577	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
50) Toluene-d8	10.08	98	683794	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
62) 4-Bromofluorobenzene	12.41	95	222688	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
16) Acetone	3.80	43	17194	7.134	ug/l	92
18) Methyl Acetate	4.31	43	6799	1.241	ug/l	96
20) Methylene Chloride	4.53	84	51186	12.438	ug/l	97
43) Isopropyl Acetate	8.15	43	81201	8.603	ug/l #	90
95) Naphthalene	15.13	128	13227	3.390	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102419\
Data File : VN058927.D
Acq On : 23 Oct 2019 14:18
Operator : JC/SP
Sample : K5139-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-03-102119

Quant Time: Oct 24 01:59:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

