

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102419\
 Data File : VN058939.D
 Acq On : 23 Oct 2019 18:42
 Operator : JC/SP
 Sample : K5508-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S32-0668

Quant Time: Oct 24 02:32:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	299685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	510281	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	457690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	176206	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	244470	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
35) Dibromofluoromethane	7.57	113	173347	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
50) Toluene-d8	10.08	98	612156	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
62) 4-Bromofluorobenzene	12.40	95	197050	41.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.40%	
Target Compounds						
2) Dichlorodifluoromethane	1.84	85	8279	2.216	ug/l	97
4) Vinyl Chloride	2.17	62	5420	1.180	ug/l	99
16) Acetone	3.80	43	5367	2.471	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	2848	0.764	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102419\
Data File : VN058939.D
Acq On : 23 Oct 2019 18:42
Operator : JC/SP
Sample : K5508-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S32-0668

Quant Time: Oct 24 02:32:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

