

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102419\
 Data File : VN058941.D
 Acq On : 23 Oct 2019 19:26
 Operator : JC/SP
 Sample : K5508-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S32-ER-2019-2

Quant Time: Oct 24 02:37:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	303430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	507283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	455813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	170890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	251681	56.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.10%	
35) Dibromofluoromethane	7.57	113	175649	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
50) Toluene-d8	10.08	98	619779	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	12.40	95	202162	42.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.04%	
Target Compounds						
20) Methylene Chloride	4.53	84	5983	1.593	ug/l	93
29) Tetrahydrofuran	7.19	42	35937	20.579	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102419\
Data File : VN058941.D
Acq On : 23 Oct 2019 19:26
Operator : JC/SP
Sample : K5508-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S32-ER-2019-2

Quant Time: Oct 24 02:37:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

