

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102422\
 Data File : VN074997.D
 Acq On : 24 Oct 2022 20:39
 Operator : JC\MD
 Sample : PB148483TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148483TB

Quant Time: Oct 25 06:52:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	2512947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	4456220	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	5242018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	3178090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	2957709	44.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.600%
35) Dibromofluoromethane	8.171	113	2246976	42.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.720%
50) Toluene-d8	10.571	98	10498639	51.590	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.180%
62) 4-Bromofluorobenzene	12.853	95	3691042	51.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.840%
Target Compounds						
					Qvalue	
37) Ethyl Acetate	7.571	43	123267	1.567	ug/l	98
43) Isopropyl Acetate	8.694	43	785628	6.440	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102422\
Data File : VN074997.D
Acq On : 24 Oct 2022 20:39
Operator : JC\MD
Sample : PB148483TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148483TB

Quant Time: Oct 25 06:52:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

