

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102422\
 Data File : VN074977.D
 Acq On : 24 Oct 2022 12:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 06:38:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.782	0.636	18.7	92	0.00
3 P	Chloromethane	1.054	0.736	30.2#	87	0.00
4 C	Vinyl Chloride	1.415	1.120	20.8#	89	0.00
5 T	Bromomethane	1.119	1.010	9.7	100	0.00
6 T	Chloroethane	1.135	0.900	20.7	93	0.00
7 T	Trichlorofluoromethane	1.509	1.206	20.1	93	0.00
8 T	Diethyl Ether	0.593	0.556	6.2	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.896	0.726	19.0	93	0.00
10 T	Methyl Iodide	1.118	1.067	4.6	97	0.00
11 T	Tert butyl alcohol	0.281	0.279	0.7	102	0.01
12 CM	1,1-Dichloroethene	0.813	0.649	20.2#	93	0.00
13 T	Acrolein	0.154	0.177	-14.9	127	0.00
14 T	Allyl chloride	1.217	1.130	7.1	98	0.00
15 T	Acrylonitrile	0.568	0.574	-1.1	101	0.00
16 T	Acetone	0.563	0.461	18.1	94	0.00
17 T	Carbon Disulfide	1.732	1.221	29.5#	85	0.00
18 T	Methyl Acetate	1.687	1.621	3.9	104	0.00
19 T	Methyl tert-butyl Ether	3.226	3.346	-3.7	105	0.00
20 T	Methylene Chloride	1.100	0.915	16.8	100	0.00
21 T	trans-1,2-Dichloroethene	0.898	0.775	13.7	98	0.00
22 T	Diisopropyl ether	2.784	2.835	-1.8	104	0.00
23 T	Vinyl Acetate	2.191	2.054	6.3	92	0.00
24 P	1,1-Dichloroethane	1.721	1.651	4.1	99	0.00
25 T	2-Butanone	0.809	0.776	4.1	99	0.00
26 T	2,2-Dichloropropane	1.627	1.449	10.9	90	0.00
27 T	cis-1,2-Dichloroethene	1.116	1.066	4.5	100	0.00
28 T	Bromochloromethane	0.717	0.738	-2.9	99	0.00
29 T	Tetrahydrofuran	0.499	0.495	0.8	100	0.00
30 C	Chloroform	1.945	1.850	4.9#	99	0.00
31 T	Cyclohexane	1.357	1.100	18.9	88	0.00
32 T	1,1,1-Trichloroethane	1.750	1.564	10.6	92	0.00
33 S	1,2-Dichloroethane-d4	1.328	1.246	6.2	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.595	0.558	6.2	96	0.00
36 T	1,1-Dichloropropene	0.749	0.646	13.8	94	0.00
37 T	Ethyl Acetate	0.883	0.848	4.0	104	0.00
38 T	Carbon Tetrachloride	0.816	0.728	10.8	92	0.00
39 T	Methylcyclohexane	0.888	0.756	14.9	88	0.00
40 TM	Benzene	2.311	2.112	8.6	96	0.00
41 T	Methacrylonitrile	0.443	0.438	1.1	103	0.00
42 TM	1,2-Dichloroethane	0.875	0.829	5.3	102	0.00
43 T	Isopropyl Acetate	1.369	1.380	-0.8	103	0.00
44 TM	Trichloroethene	0.567	0.494	12.9	95	0.00
45 C	1,2-Dichloropropane	0.596	0.564	5.4#	100	0.00
46 T	Dibromomethane	0.436	0.430	1.4	102	0.00
47 T	Bromodichloromethane	0.880	0.855	2.8	99	0.00
48 T	Methyl methacrylate	0.599	0.622	-3.8	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102422\
 Data File : VN074977.D
 Acq On : 24 Oct 2022 12:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 06:38:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 21 09:37:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.017	0.017	0.0	97	0.00
50 S	Toluene-d8	2.283	2.127	6.8	92	0.00
51 T	4-Methyl-2-Pentanone	0.907	0.925	-2.0	101	0.00
52 CM	Toluene	1.508	1.443	4.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.868	0.960	-10.6	105	0.00
54 T	cis-1,3-Dichloropropene	0.960	0.957	0.3	101	0.00
55 T	1,1,2-Trichloroethane	0.633	0.637	-0.6	100	0.00
56 T	Ethyl methacrylate	0.953	1.017	-6.7	105	0.00
57 T	1,3-Dichloropropane	1.074	1.066	0.7	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.381	0.402	-5.5	95	0.00
59 T	2-Hexanone	0.688	0.705	-2.5	100	0.00
60 T	Dibromochloromethane	0.664	0.709	-6.8	105	0.00
61 T	1,2-Dibromoethane	0.638	0.664	-4.1	104	0.00
62 S	4-Bromofluorobenzene	0.805	0.811	-0.7	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.409	0.343	16.1	97	0.00
65 PM	Chlorobenzene	1.512	1.405	7.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.570	0.571	-0.2	104	0.00
67 C	Ethyl Benzene	2.684	2.494	7.1#	96	0.00
68 T	m/p-Xylenes	1.047	1.002	4.3	98	0.00
69 T	o-Xylene	1.048	1.026	2.1	99	0.00
70 T	Styrene	1.662	1.710	-2.9	99	0.00
71 P	Bromoform	0.430	0.449	-4.4	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.084	3.636	11.0	99	0.00
74 T	N-amyl acetate	1.641	1.522	7.3	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.525	1.372	10.0	105	0.00
76 T	1,2,3-Trichloropropane	1.356	1.193	12.0	103	0.00
77 T	Bromobenzene	0.937	0.821	12.4	102	0.00
78 T	n-propylbenzene	4.780	4.254	11.0	98	0.00
79 T	2-Chlorotoluene	2.961	2.624	11.4	101	0.00
80 T	1,3,5-Trimethylbenzene	3.528	3.231	8.4	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.423	0.466	-10.2	109	0.00
82 T	4-Chlorotoluene	2.961	2.624	11.4	101	0.00
83 T	tert-Butylbenzene	3.122	2.811	10.0	96	0.00
84 T	1,2,4-Trimethylbenzene	3.563	3.313	7.0	99	0.00
85 T	sec-Butylbenzene	4.383	4.041	7.8	96	0.00
86 T	p-Isopropyltoluene	3.610	3.405	5.7	97	0.00
87 T	1,3-Dichlorobenzene	1.792	1.710	4.6	101	0.00
88 T	1,4-Dichlorobenzene	1.817	1.734	4.6	103	0.00
89 T	n-Butylbenzene	3.083	2.862	7.2	94	0.00
90 T	Hexachloroethane	0.699	0.639	8.6	100	0.00
91 T	1,2-Dichlorobenzene	1.837	1.741	5.2	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.317	0.308	2.8	100	0.00
93 T	1,2,4-Trichlorobenzene	0.799	0.808	-1.1	105	0.00
94 T	Hexachlorobutadiene	0.413	0.354	14.3	97	0.00
95 T	Naphthalene	3.038	3.213	-5.8	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102422\
Data File : VN074977.D
Acq On : 24 Oct 2022 12:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 25 06:38:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102022W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 21 09:37:47 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.804	0.794	1.2	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6