

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
 Data File : VN079662.D
 Acq On : 24 Oct 2023 17:14
 Operator : JC\MD
 Sample : 05029-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 231010056-08-TRIP-BLANK

Quant Time: Oct 25 02:00:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	244177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	470040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	467991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	150719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	201179	50.814	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.620%
35) Dibromofluoromethane	8.171	113	180752	53.511	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.020%
50) Toluene-d8	10.571	98	559968	45.816	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.640%
62) 4-Bromofluorobenzene	12.853	95	194593	49.612	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
Data File : VN079662.D
Acq On : 24 Oct 2023 17:14
Operator : JC\MD
Sample : 05029-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
231010056-08-TRIP-BLANK

Quant Time: Oct 25 02:00:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

