

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
 Data File : VN079666.D
 Acq On : 24 Oct 2023 18:49
 Operator : JC\MD
 Sample : 05029-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 231010082-43

Quant Time: Oct 25 02:05:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	246849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	479603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	470653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	150014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	207166	51.760	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.520%
35) Dibromofluoromethane	8.171	113	185081	53.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.400%
50) Toluene-d8	10.570	98	570182	45.721	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.440%
62) 4-Bromofluorobenzene	12.853	95	193546	48.361	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.720%
Target Compounds						
60) Dibromochloromethane	11.359	129	2330	0.639	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
Data File : VN079666.D
Acq On : 24 Oct 2023 18:49
Operator : JC\MD
Sample : 05029-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
231010082-43

Quant Time: Oct 25 02:05:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

