

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
 Data File : VN079655.D
 Acq On : 24 Oct 2023 13:24
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 01:36:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.716	0.645	9.9	88	0.00
3 P	Chloromethane	0.894	0.723	19.1	84	0.00
4 C	Vinyl Chloride	0.930	0.828	11.0#	90	0.00
5 T	Bromomethane	0.557	0.530	4.8	104	0.00
6 T	Chloroethane	0.701	0.576	17.8	92	0.00
7 T	Trichlorofluoromethane	1.098	1.036	5.6	99	0.00
8 T	Diethyl Ether	0.391	0.396	-1.3	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.642	0.601	6.4	103	0.00
10 T	Methyl Iodide	0.557	0.664	-19.2	106	0.00
11 T	Tert butyl alcohol	0.122	0.115	5.7	98	0.00
12 CM	1,1-Dichloroethene	0.591	0.563	4.7#	98	0.00
13 T	Acrolein	0.121	0.155	-28.1#	127	0.00
14 T	Allyl chloride	0.953	0.896	6.0	97	0.00
15 T	Acrylonitrile	0.354	0.330	6.8	92	0.00
16 T	Acetone	0.322	0.329	-2.2	102	0.00
17 T	Carbon Disulfide	1.843	1.505	18.3	87	0.00
18 T	Methyl Acetate	1.574	1.389	11.8	94	0.00
19 T	Methyl tert-butyl Ether	1.919	2.008	-4.6	102	0.00
20 T	Methylene Chloride	0.816	0.692	15.2	101	0.00
21 T	trans-1,2-Dichloroethene	0.690	0.626	9.3	98	0.00
22 T	Diisopropyl ether	2.138	2.248	-5.1	103	0.00
23 T	Vinyl Acetate	1.202	1.215	-1.1	94	0.00
24 P	1,1-Dichloroethane	1.308	1.286	1.7	102	0.00
25 T	2-Butanone	0.477	0.463	2.9	93	0.00
26 T	2,2-Dichloropropane	1.000	1.073	-7.3	107	0.00
27 T	cis-1,2-Dichloroethene	0.750	0.761	-1.5	104	0.00
28 T	Bromochloromethane	0.622	0.601	3.4	94	0.00
29 T	Tetrahydrofuran	0.309	0.292	5.5	91	0.00
30 C	Chloroform	1.322	1.313	0.7#	104	0.00
31 T	Cyclohexane	1.150	1.009	12.3	95	0.00
32 T	1,1,1-Trichloroethane	1.112	1.110	0.2	103	0.00
33 S	1,2-Dichloroethane-d4	0.811	0.784	3.3	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.359	0.364	-1.4	104	0.00
36 T	1,1-Dichloropropene	0.539	0.561	-4.1	101	0.00
37 T	Ethyl Acetate	0.537	0.536	0.2	93	-0.01
38 T	Carbon Tetrachloride	0.540	0.555	-2.8	101	0.00
39 T	Methylcyclohexane	0.518	0.540	-4.2	96	0.00
40 TM	Benzene	1.602	1.641	-2.4	101	0.00
41 T	Methacrylonitrile	0.282	0.282	0.0	91	0.00
42 TM	1,2-Dichloroethane	0.582	0.595	-2.2	103	0.00
43 T	Isopropyl Acetate	0.940	0.910	3.2	96	0.00
44 TM	Trichloroethene	0.377	0.386	-2.4	103	0.00
45 C	1,2-Dichloropropane	0.422	0.439	-4.0#	104	0.00
46 T	Dibromomethane	0.273	0.286	-4.8	101	0.00
47 T	Bromodichloromethane	0.562	0.605	-7.7	104	0.00
48 T	Methyl methacrylate	0.403	0.435	-7.9	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
 Data File : VN079655.D
 Acq On : 24 Oct 2023 13:24
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 01:36:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	84	0.00
50 S	Toluene-d8	1.300	1.378	-6.0	102	0.00
51 T	4-Methyl-2-Pentanone	0.531	0.543	-2.3	94	0.00
52 CM	Toluene	0.943	1.010	-7.1#	100	0.00
53 T	t-1,3-Dichloropropene	0.567	0.647	-14.1	105	0.00
54 T	cis-1,3-Dichloropropene	0.620	0.697	-12.4	104	0.00
55 T	1,1,2-Trichloroethane	0.378	0.399	-5.6	103	0.00
56 T	Ethyl methacrylate	0.520	0.604	-16.2	99	0.00
57 T	1,3-Dichloropropane	0.654	0.713	-9.0	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.274	-37.7#	119	0.00
59 T	2-Hexanone	0.383	0.403	-5.2	93	0.00
60 T	Dibromochloromethane	0.380	0.430	-13.2	106	0.00
61 T	1,2-Dibromoethane	0.376	0.398	-5.9	103	0.00
62 S	4-Bromofluorobenzene	0.417	0.461	-10.6	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.344	0.339	1.5	101	0.00
65 PM	Chlorobenzene	1.114	1.165	-4.6	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.408	0.435	-6.6	106	0.00
67 C	Ethyl Benzene	1.935	2.145	-10.9#	102	0.00
68 T	m/p-Xylenes	0.703	0.800	-13.8	103	0.00
69 T	o-Xylene	0.692	0.769	-11.1	102	0.00
70 T	Styrene	1.092	1.285	-17.7	102	0.00
71 P	Bromoform	0.298	0.310	-4.0	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.652	4.583	1.5	102	0.00
74 T	N-amyl acetate	1.907	1.892	0.8	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.814	1.484	18.2	99	0.00
76 T	1,2,3-Trichloropropane	1.442	1.320	8.5	96	0.00
77 T	Bromobenzene	1.141	1.054	7.6	102	0.00
78 T	n-propylbenzene	5.263	5.474	-4.0	102	0.00
79 T	2-Chlorotoluene	3.368	3.299	2.0	105	0.00
80 T	1,3,5-Trimethylbenzene	3.668	3.767	-2.7	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.497	0.453	8.9	94	0.00
82 T	4-Chlorotoluene	3.205	3.243	-1.2	103	0.00
83 T	tert-Butylbenzene	3.005	3.190	-6.2	104	0.00
84 T	1,2,4-Trimethylbenzene	3.501	3.802	-8.6	104	0.00
85 T	sec-Butylbenzene	4.192	4.369	-4.2	102	0.00
86 T	p-Isopropyltoluene	3.216	3.568	-10.9	103	0.00
87 T	1,3-Dichlorobenzene	1.809	1.881	-4.0	103	0.00
88 T	1,4-Dichlorobenzene	1.868	1.839	1.6	102	0.00
89 T	n-Butylbenzene	2.568	2.954	-15.0	102	0.00
90 T	Hexachloroethane	0.756	0.682	9.8	102	0.00
91 T	1,2-Dichlorobenzene	1.765	1.782	-1.0	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.235	12.6	90	0.00
93 T	1,2,4-Trichlorobenzene	0.547	0.727	-32.9#	109	0.00
94 T	Hexachlorobutadiene	0.426	0.530	-24.4	131	0.00
95 T	Naphthalene	1.669	1.860	-11.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102423\
Data File : VN079655.D
Acq On : 24 Oct 2023 13:24
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 25 01:36:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.652	0.774	-18.7	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6