

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102425\
 Data File : VN088095.D
 Acq On : 24 Oct 2025 15:39
 Operator : JC\MD
 Sample : Q3373-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Oct 25 01:06:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 25 00:15:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	214510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	479905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	453957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	217438	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	219158	59.083	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.160%
35) Dibromofluoromethane	8.147	113	166962	51.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.580%
50) Toluene-d8	10.547	98	615294	49.530	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.829	95	228168	52.007	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102425\
Data File : VN088095.D
Acq On : 24 Oct 2025 15:39
Operator : JC\MD
Sample : Q3373-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Oct 25 01:06:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 25 00:15:22 2025
Response via : Initial Calibration

