

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102518\
 Data File : VN052004.D
 Acq On : 25 Oct 2018 11:47
 Operator : MD\SY
 Sample : PB114157ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114157ZHE#01

Quant Time: Oct 26 01:54:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1134355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1652540	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1592397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	762030	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	589468	58.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.70%	
35) Dibromofluoromethane	7.59	113	608786	57.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.46%	
50) Toluene-d8	10.09	98	2276308	59.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.48%	
62) 4-Bromofluorobenzene	12.40	95	752659	57.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.92%	
Target Compounds						
16) Acetone	3.82	43	9181	5.516	ug/l #	79
43) Isopropyl Acetate	8.16	43	326365	30.742	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102518\
Data File : VN052004.D
Acq On : 25 Oct 2018 11:47
Operator : MD\SY
Sample : PB114157ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114157ZHE#01

Quant Time: Oct 26 01:54:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

