

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102518\
 Data File : VN052010.D
 Acq On : 25 Oct 2018 14:28
 Operator : MD\SY
 Sample : PB114155TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114155TB

Manual Integrations
 APPROVED

MMDadoda
 10/26/2018 9:19:42 AM

Quant Time: Oct 26 02:12:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1139212	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1660893	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1565898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	699414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	587613	57.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.84%	
35) Dibromofluoromethane	7.59	113	606905	57.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.52%	
50) Toluene-d8	10.09	98	2275197	59.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.82%	
62) 4-Bromofluorobenzene	12.40	95	707288	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
Target Compounds						
16) Acetone	3.82	43	8423m	5.039	ug/l	
43) Isopropyl Acetate	8.16	43	315030	29.525	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102518\
Data File : VN052010.D
Acq On : 25 Oct 2018 14:28
Operator : MD\SY
Sample : PB114155TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114155TB

Manual Integrations
APPROVED

MMDadoda
10/26/2018 9:19:42 AM

Quant Time: Oct 26 02:12:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

