

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102518\
 Data File : VN052013.D
 Acq On : 25 Oct 2018 16:30
 Operator : MD\SY
 Sample : VN1025WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1025WBL02

Quant Time: Oct 26 02:21:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	166567	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	920044	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	796350	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	291909	28.80	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.00%
60) 4-Bromofluorobenzene	12.40	95	303283	26.41	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.03%
63) Toluene-d8	10.09	98	1029579	29.77	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.23%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102518\
Data File : VN052013.D
Acq On : 25 Oct 2018 16:30
Operator : MD\SY
Sample : VN1025WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1025WBL02

Quant Time: Oct 26 02:21:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

