

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102519\
 Data File : VN058950.D
 Acq On : 24 Oct 2019 11:10
 Operator : JC/SP
 Sample : VN1025WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1025WBL02

Quant Time: Oct 25 01:45:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	346867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	552220	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	499631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	187779	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	267494	52.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.16%	
35) Dibromofluoromethane	7.57	113	189230	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
50) Toluene-d8	10.08	98	708525	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.40	95	226231	43.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102519\
Data File : VN058950.D
Acq On : 24 Oct 2019 11:10
Operator : JC/SP
Sample : VN1025WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1025WBL02

Quant Time: Oct 25 01:45:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

