

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102519\
 Data File : VN058958.D
 Acq On : 24 Oct 2019 14:06
 Operator : JC/SP
 Sample : PB124243TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124243TB

Quant Time: Oct 25 02:08:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	340252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	551843	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	515523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	184099	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	253397	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
35) Dibromofluoromethane	7.57	113	181932	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.08	98	709482	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
62) 4-Bromofluorobenzene	12.40	95	227248	43.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.88%	
Target Compounds						
20) Methylene Chloride	4.53	84	13356	3.172	ug/l	95
43) Isopropyl Acetate	8.15	43	107734	10.912	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102519\
Data File : VN058958.D
Acq On : 24 Oct 2019 14:06
Operator : JC/SP
Sample : PB124243TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124243TB

Quant Time: Oct 25 02:08:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

