

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102519\
 Data File : VN058971.D
 Acq On : 24 Oct 2019 18:52
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 03:16:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.623	0.590	5.3	78	0.00
3 P	Chloromethane	0.772	0.709	8.2	78	0.00
4 C	Vinyl Chloride	0.767	0.743	3.1#	81	0.00
5 T	Bromomethane	0.480	0.400	16.7	69	0.00
6 T	Chloroethane	0.492	0.464	5.7	80	0.00
7 T	Trichlorofluoromethane	0.949	0.907	4.4	78	0.00
8 T	Diethyl Ether	0.328	0.332	-1.2	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.526	0.509	3.2	83	0.00
10 T	Methyl Iodide	0.692	0.575	16.9	69	0.00
11 T	Tert butyl alcohol	0.122	0.116	4.9	82	0.00
12 CM	1,1-Dichloroethene	0.493	0.485	1.6#	82	0.00
13 T	Acrolein	0.084	0.044	47.6#	42#	0.00
14 T	Allyl chloride	0.997	1.001	-0.4	83	0.00
15 T	Acrylonitrile	0.306	0.324	-5.9	85	0.00
16 T	Acetone	0.362	0.318	12.2	69	0.00
17 T	Carbon Disulfide	1.588	1.418	10.7	74	0.00
18 T	Methyl Acetate	0.824	0.851	-3.3	90	0.00
19 T	Methyl tert-butyl Ether	1.726	1.815	-5.2	85	0.00
20 T	Methylene Chloride	0.619	0.597	3.6	86	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.542	0.4	82	0.00
22 T	Diisopropyl ether	1.951	2.107	-8.0	86	0.00
23 T	Vinyl Acetate	1.567	1.777	-13.4	84	0.00
24 P	1,1-Dichloroethane	1.131	1.132	-0.1	85	0.00
25 T	2-Butanone	0.470	0.464	1.3	78	0.00
26 T	2,2-Dichloropropane	0.978	0.950	2.9	81	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.632	-1.6	83	0.00
28 T	Bromochloromethane	0.377	0.376	0.3	83	0.00
29 T	Tetrahydrofuran	0.288	0.306	-6.3	84	0.00
30 C	Chloroform	1.135	1.129	0.5#	85	0.00
31 T	Cyclohexane	1.054	1.002	4.9	80	0.00
32 T	1,1,1-Trichloroethane	0.956	0.967	-1.2	83	0.00
33 S	1,2-Dichloroethane-d4	0.733	0.698	4.8	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.329	0.326	0.9	81	0.00
36 T	1,1-Dichloropropene	0.512	0.527	-2.9	82	0.00
37 T	Ethyl Acetate	0.558	0.587	-5.2	81	0.00
38 T	Carbon Tetrachloride	0.527	0.531	-0.8	83	0.00
39 T	Methylcyclohexane	0.575	0.592	-3.0	80	0.00
40 TM	Benzene	1.503	1.556	-3.5	83	0.00
41 T	Methacrylonitrile	0.248	0.255	-2.8	86	0.00
42 TM	1,2-Dichloroethane	0.579	0.579	0.0	82	0.00
43 T	Isopropyl Acetate	0.895	0.963	-7.6	85	0.00
44 TM	Trichloroethene	0.366	0.368	-0.5	82	0.00
45 C	1,2-Dichloropropane	0.412	0.427	-3.6#	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102519\
 Data File : VN058971.D
 Acq On : 24 Oct 2019 18:52
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 03:16:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.268	-2.7	84	0.00
47 T	Bromodichloromethane	0.537	0.562	-4.7	85	0.00
48 T	Methyl methacrylate	0.406	0.443	-9.1	83	0.00
49 T	1,4-Dioxane	0.007	0.006	14.3	78	0.00
50 S	Toluene-d8	1.276	1.272	0.3	81	0.00
51 T	4-Methyl-2-Pentanone	0.522	0.612	-17.2	86	0.00
52 CM	Toluene	0.891	0.969	-8.8#	84	0.00
53 T	t-1,3-Dichloropropene	0.595	0.639	-7.4	84	0.00
54 T	cis-1,3-Dichloropropene	0.635	0.669	-5.4	82	0.00
55 T	1,1,2-Trichloroethane	0.367	0.392	-6.8	88	0.00
56 T	Ethyl methacrylate	0.540	0.615	-13.9	85	0.00
57 T	1,3-Dichloropropane	0.645	0.683	-5.9	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.270	-15.4	86	0.00
59 T	2-Hexanone	0.404	0.456	-12.9	82	0.00
60 T	Dibromochloromethane	0.395	0.423	-7.1	85	0.00
61 T	1,2-Dibromoethane	0.376	0.402	-6.9	85	0.00
62 S	4-Bromofluorobenzene	0.469	0.470	-0.2	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.350	0.348	0.6	82	0.00
65 PM	Chlorobenzene	1.018	1.066	-4.7	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.410	-6.8	87	0.00
67 C	Ethyl Benzene	1.786	1.974	-10.5#	84	0.00
68 T	m/p-Xylenes	0.662	0.725	-9.5	83	0.00
69 T	o-Xylene	0.632	0.686	-8.5	83	0.00
70 T	Styrene	1.030	1.179	-14.5	85	0.00
71 P	Bromoform	0.298	0.322	-8.1	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.654	3.853	-5.4	85	0.00
74 T	N-amyl acetate	1.657	1.805	-8.9	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.301	-1.2	86	0.00
76 T	1,2,3-Trichloropropane	1.159	1.117	3.6	85	0.00
77 T	Bromobenzene	0.901	0.912	-1.2	85	0.00
78 T	n-propylbenzene	4.319	4.583	-6.1	84	0.00
79 T	2-Chlorotoluene	2.644	2.712	-2.6	85	0.00
80 T	1,3,5-Trimethylbenzene	3.070	3.309	-7.8	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.439	1.3	80	0.00
82 T	4-Chlorotoluene	2.702	2.809	-4.0	84	0.00
83 T	tert-Butylbenzene	2.625	2.778	-5.8	85	0.00
84 T	1,2,4-Trimethylbenzene	3.028	3.299	-8.9	83	0.00
85 T	sec-Butylbenzene	3.500	3.764	-7.5	83	0.00
86 T	p-Isopropyltoluene	3.088	3.395	-9.9	83	0.00
87 T	1,3-Dichlorobenzene	1.610	1.658	-3.0	85	0.00
88 T	1,4-Dichlorobenzene	1.612	1.633	-1.3	83	0.00
89 T	n-Butylbenzene	2.837	3.059	-7.8	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102519\
Data File : VN058971.D
Acq On : 24 Oct 2019 18:52
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 25 03:16:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.627	-0.2	85	0.00
91 T	1,2-Dichlorobenzene	1.583	1.641	-3.7	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.291	0.277	4.8	82	0.00
93 T	1,2,4-Trichlorobenzene	0.873	0.958	-9.7	84	0.00
94 T	Hexachlorobutadiene	0.513	0.502	2.1	82	0.00
95 T	Naphthalene	2.436	2.765	-13.5	83	0.00
96 T	1,2,3-Trichlorobenzene	0.858	0.948	-10.5	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6