

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102522\
 Data File : VN075009.D
 Acq On : 25 Oct 2022 15:26
 Operator : JC\MD
 Sample : VN1025WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1025WBL01

Quant Time: Oct 27 10:25:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102522W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 27 10:19:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	13676	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	78300	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	66204	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	41092	32.861	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	109.533%
60) 4-Bromofluorobenzene	12.847	95	34265	26.341	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.800%
63) Toluene-d8	10.571	98	118661	32.046	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	106.833%
Target Compounds						
12) Methyl Acetate	5.053	43	334	0.252	ug/l	Qvalue # 49

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102522\
Data File : VN075009.D
Acq On : 25 Oct 2022 15:26
Operator : JC\MD
Sample : VN1025WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1025WBL01

Quant Time: Oct 27 10:25:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N102522W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 27 10:19:47 2022
Response via : Initial Calibration

