

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102618\
 Data File : VN052020.D
 Acq On : 26 Oct 2018 11:29
 Operator : MD\SY
 Sample : VN1026WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1026WBL01

Quant Time: Oct 26 23:56:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1210733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1725125	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1612656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	727479	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	594740	55.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.30%	
35) Dibromofluoromethane	7.59	113	635172	57.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.40%	
50) Toluene-d8	10.09	98	2331259	58.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.20%	
62) 4-Bromofluorobenzene	12.40	95	734207	53.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102618\
Data File : VN052020.D
Acq On : 26 Oct 2018 11:29
Operator : MD\SY
Sample : VN1026WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1026WBL01

Quant Time: Oct 26 23:56:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

