

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102618\
 Data File : VN052026.D
 Acq On : 26 Oct 2018 14:20
 Operator : MD\SY
 Sample : PB114195ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114195ZHE#03

Quant Time: Oct 27 00:21:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1193137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1728573	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1664141	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	788700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	609577	57.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.72%	
35) Dibromofluoromethane	7.59	113	645684	58.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.08%	
50) Toluene-d8	10.09	98	2380529	59.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.44%	
62) 4-Bromofluorobenzene	12.40	95	764932	55.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.66%	
Target Compounds						
16) Acetone	3.83	43	8927	5.099	ug/l	88
20) Methylene Chloride	4.56	84	13015	1.267	ug/l	92
43) Isopropyl Acetate	8.17	43	348510	31.384	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102618\
 Data File : VN052026.D
 Acq On : 26 Oct 2018 14:20
 Operator : MD\SY
 Sample : PB114195ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114195ZHE#03

Quant Time: Oct 27 00:21:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

