

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102618\
 Data File : VN052035.D
 Acq On : 26 Oct 2018 18:20
 Operator : MD\SY
 Sample : PB114195ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114195ZHE#12

Quant Time: Oct 27 00:46:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1221193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1783223	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1678188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	721917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	620940	57.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.18%	
35) Dibromofluoromethane	7.59	113	663214	58.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.56%	
50) Toluene-d8	10.09	98	2440046	59.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.68%	
62) 4-Bromofluorobenzene	12.40	95	750301	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
Target Compounds						
16) Acetone	3.82	43	32307	18.029	ug/l	95
20) Methylene Chloride	4.56	84	13796	1.312	ug/l	92
43) Isopropyl Acetate	8.17	43	337024	29.420	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102618\
Data File : VN052035.D
Acq On : 26 Oct 2018 18:20
Operator : MD\SY
Sample : PB114195ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114195ZHE#12

Quant Time: Oct 27 00:46:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

