

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102618\
 Data File : VN052038.D
 Acq On : 26 Oct 2018 19:40
 Operator : MD\SY
 Sample : J5675-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Oct 27 00:52:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 10:12:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1246203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1842470	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1760485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	827410	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	643006	57.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.86%	
35) Dibromofluoromethane	7.59	113	686405	58.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.76%	
50) Toluene-d8	10.09	98	2507372	59.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.04%	
62) 4-Bromofluorobenzene	12.40	95	808120	55.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.68%	
Target Compounds						
8) Diethyl Ether	3.42	74	6375	1.261	ug/l	74
16) Acetone	3.82	43	37293	20.394	ug/l	95
20) Methylene Chloride	4.55	84	22253	2.074	ug/l	96
43) Isopropyl Acetate	8.16	43	322780	27.270	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102618\
Data File : VN052038.D
Acq On : 26 Oct 2018 19:40
Operator : MD\SY
Sample : J5675-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-13

Quant Time: Oct 27 00:52:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 10:12:55 2018
Response via : Initial Calibration

