


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN102618\
Data File : VN052018.D
Acq On : 26 Oct 2018 9:52
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102318W.M
Title : SW846 8260
Last Update : Wed Oct 24 10:12:55 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	18770	PASS
75	95	30	60	48.2	54186	PASS
95	95	100	100	100.0	112522	PASS
96	95	5	9	6.8	7689	PASS
173	174	0.00	2	0.9	989	PASS
174	95	50	100	99.4	111896	PASS
175	174	5	9	7.7	8625	PASS
176	174	95	101	97.7	109314	PASS
177	176	5	9	6.8	7439	PASS