

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102619\
 Data File : VN058975.D
 Acq On : 25 Oct 2019 10:42
 Operator : JC/SP
 Sample : VN1026WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1026WBL01

Quant Time: Oct 25 23:30:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	350367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	556059	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	492519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	181870	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	266892	51.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.88%	
35) Dibromofluoromethane	7.57	113	191471	52.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.74%	
50) Toluene-d8	10.08	98	711982	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.40	95	225128	43.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102619\
Data File : VN058975.D
Acq On : 25 Oct 2019 10:42
Operator : JC/SP
Sample : VN1026WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1026WBL01

Quant Time: Oct 25 23:30:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

