

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102619\
 Data File : VN058976.D
 Acq On : 25 Oct 2019 11:04
 Operator : JC/SP
 Sample : K5542-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIPMENT-BLANK-102419

Quant Time: Oct 25 23:32:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	333726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	522429	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	455694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	151062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	244944	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	7.57	113	175102	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
50) Toluene-d8	10.08	98	650131	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
62) 4-Bromofluorobenzene	12.40	95	193148	39.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.90%	
Target Compounds						
8) Diethyl Ether	3.39	74	19268	8.807	ug/l	95
16) Acetone	3.79	43	81575	33.728	ug/l	100
20) Methylene Chloride	4.53	84	472151	114.326	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102619\
Data File : VN058976.D
Acq On : 25 Oct 2019 11:04
Operator : JC/SP
Sample : K5542-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIPMENT-BLANK-102419

Quant Time: Oct 25 23:32:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

