

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102619\
 Data File : VN058973.D
 Acq On : 25 Oct 2019 9:08
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 23:11:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.623	0.594	4.7	82	0.00
3 P	Chloromethane	0.772	0.730	5.4	84	0.00
4 C	Vinyl Chloride	0.767	0.757	1.3#	86	0.00
5 T	Bromomethane	0.480	0.440	8.3	80	0.00
6 T	Chloroethane	0.492	0.470	4.5	85	0.00
7 T	Trichlorofluoromethane	0.949	0.939	1.1	85	0.00
8 T	Diethyl Ether	0.328	0.333	-1.5	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.526	0.513	2.5	87	0.00
10 T	Methyl Iodide	0.692	0.586	15.3	73	0.00
11 T	Tert butyl alcohol	0.122	0.119	2.5	87	0.00
12 CM	1,1-Dichloroethene	0.493	0.485	1.6#	86	0.00
13 T	Acrolein	0.084	0.099	-17.9	99	0.00
14 T	Allyl chloride	0.997	1.014	-1.7	88	0.00
15 T	Acrylonitrile	0.306	0.324	-5.9	88	0.00
16 T	Acetone	0.362	0.400	-10.5	90	0.00
17 T	Carbon Disulfide	1.588	1.457	8.2	80	0.00
18 T	Methyl Acetate	0.824	0.837	-1.6	93	0.00
19 T	Methyl tert-butyl Ether	1.726	1.818	-5.3	89	0.00
20 T	Methylene Chloride	0.619	0.596	3.7	89	0.00
21 T	trans-1,2-Dichloroethene	0.544	0.547	-0.6	87	0.00
22 T	Diisopropyl ether	1.951	2.106	-7.9	90	0.00
23 T	Vinyl Acetate	1.567	1.771	-13.0	88	0.00
24 P	1,1-Dichloroethane	1.131	1.133	-0.2	88	0.00
25 T	2-Butanone	0.470	0.495	-5.3	87	0.00
26 T	2,2-Dichloropropane	0.978	0.999	-2.1	89	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.631	-1.4	87	0.00
28 T	Bromochloromethane	0.377	0.388	-2.9	89	0.00
29 T	Tetrahydrofuran	0.288	0.305	-5.9	87	0.00
30 C	Chloroform	1.135	1.123	1.1#	88	0.00
31 T	Cyclohexane	1.054	1.001	5.0	84	0.00
32 T	1,1,1-Trichloroethane	0.956	0.957	-0.1	86	0.00
33 S	1,2-Dichloroethane-d4	0.733	0.741	-1.1	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.329	0.347	-5.5	89	0.00
36 T	1,1-Dichloropropene	0.512	0.535	-4.5	85	0.00
37 T	Ethyl Acetate	0.558	0.598	-7.2	85	0.00
38 T	Carbon Tetrachloride	0.527	0.532	-0.9	85	0.00
39 T	Methylcyclohexane	0.575	0.597	-3.8	83	0.00
40 TM	Benzene	1.503	1.592	-5.9	87	0.00
41 T	Methacrylonitrile	0.248	0.280	-12.9	97	0.00
42 TM	1,2-Dichloroethane	0.579	0.590	-1.9	86	0.00
43 T	Isopropyl Acetate	0.895	0.962	-7.5	87	0.00
44 TM	Trichloroethene	0.366	0.379	-3.6	87	0.00
45 C	1,2-Dichloropropane	0.412	0.443	-7.5#	89	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102619\
 Data File : VN058973.D
 Acq On : 25 Oct 2019 9:08
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 23:11:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.270	-3.4	87	0.00
47 T	Bromodichloromethane	0.537	0.576	-7.3	89	0.00
48 T	Methyl methacrylate	0.406	0.451	-11.1	87	0.00
49 T	1,4-Dioxane	0.007	0.008	-14.3	100	0.00
50 S	Toluene-d8	1.276	1.354	-6.1	89	0.00
51 T	4-Methyl-2-Pentanone	0.522	0.610	-16.9	88	0.00
52 CM	Toluene	0.891	0.991	-11.2#	89	0.00
53 T	t-1,3-Dichloropropene	0.595	0.653	-9.7	88	0.00
54 T	cis-1,3-Dichloropropene	0.635	0.704	-10.9	89	0.00
55 T	1,1,2-Trichloroethane	0.367	0.394	-7.4	91	0.00
56 T	Ethyl methacrylate	0.540	0.624	-15.6	89	0.00
57 T	1,3-Dichloropropane	0.645	0.693	-7.4	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.267	-14.1	88	0.00
59 T	2-Hexanone	0.404	0.476	-17.8	88	0.00
60 T	Dibromochloromethane	0.395	0.433	-9.6	90	0.00
61 T	1,2-Dibromoethane	0.376	0.406	-8.0	89	0.00
62 S	4-Bromofluorobenzene	0.469	0.506	-7.9	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.350	0.361	-3.1	88	0.00
65 PM	Chlorobenzene	1.018	1.080	-6.1	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.408	-6.2	89	0.00
67 C	Ethyl Benzene	1.786	2.004	-12.2#	88	0.00
68 T	m/p-Xylenes	0.662	0.745	-12.5	88	0.00
69 T	o-Xylene	0.632	0.701	-10.9	88	0.00
70 T	Styrene	1.030	1.193	-15.8	89	0.00
71 P	Bromoform	0.298	0.324	-8.7	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.654	3.879	-6.2	88	0.00
74 T	N-amyl acetate	1.657	1.788	-7.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.284	0.1	88	0.00
76 T	1,2,3-Trichloropropane	1.159	1.111	4.1	87	0.00
77 T	Bromobenzene	0.901	0.936	-3.9	90	0.00
78 T	n-propylbenzene	4.319	4.737	-9.7	90	0.00
79 T	2-Chlorotoluene	2.644	2.759	-4.3	89	0.00
80 T	1,3,5-Trimethylbenzene	3.070	3.350	-9.1	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.464	-4.3	88	0.00
82 T	4-Chlorotoluene	2.702	2.879	-6.6	89	0.00
83 T	tert-Butylbenzene	2.625	2.760	-5.1	87	0.00
84 T	1,2,4-Trimethylbenzene	3.028	3.385	-11.8	88	0.00
85 T	sec-Butylbenzene	3.500	3.830	-9.4	88	0.00
86 T	p-Isopropyltoluene	3.088	3.485	-12.9	89	0.00
87 T	1,3-Dichlorobenzene	1.610	1.717	-6.6	91	0.00
88 T	1,4-Dichlorobenzene	1.612	1.697	-5.3	90	0.00
89 T	n-Butylbenzene	2.837	3.156	-11.2	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102619\
Data File : VN058973.D
Acq On : 25 Oct 2019 9:08
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 25 23:11:35 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.636	-1.6	89	0.00
91 T	1,2-Dichlorobenzene	1.583	1.675	-5.8	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.291	0.271	6.9	83	0.00
93 T	1,2,4-Trichlorobenzene	0.873	1.005	-15.1	91	0.00
94 T	Hexachlorobutadiene	0.513	0.528	-2.9	89	0.00
95 T	Naphthalene	2.436	2.775	-13.9	86	0.00
96 T	1,2,3-Trichlorobenzene	0.858	0.966	-12.6	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6