

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102620\
 Data File : VN064311.D
 Acq On : 26 Oct 2020 19:28
 Operator : JC/MD
 Sample : VN1026WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1026WBL01

Quant Time: Oct 27 02:44:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102620W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Oct 26 17:44:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	49892	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	278953	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	251773	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	142789	31.81	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.03%
60) 4-Bromofluorobenzene	12.38	95	123753	27.48	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.60%
63) Toluene-d8	10.06	98	336264	29.95	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.83%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102620\
Data File : VN064311.D
Acq On : 26 Oct 2020 19:28
Operator : JC/MD
Sample : VN1026WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1026WBL01

Quant Time: Oct 27 02:44:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N102620W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Oct 26 17:44:53 2020
Response via : Initial Calibration

