

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102621\
 Data File : VN069163.D
 Acq On : 25 Oct 2021 20:15
 Operator : JC/MD
 Sample : M4307-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Oct 26 01:36:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	344968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	606871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	580937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	221003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225074	49.485	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.980%		
35) Dibromofluoromethane	8.024	113	176834	50.083	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.160%		
50) Toluene-d8	10.443	98	748575	52.987	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.980%		
62) 4-Bromofluorobenzene	12.733	95	271895	51.011	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	102.020%		
Target Compounds						Qvalue
16) Acetone	4.299	43	8087	4.765	ug/l	95
89) n-Butylbenzene	13.946	91	3766	0.305	ug/l	91
95) Naphthalene	15.509	128	15557	5.213	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102621\
Data File : VN069163.D
Acq On : 25 Oct 2021 20:15
Operator : JC/MD
Sample : M4307-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Oct 26 01:36:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

