

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102621\
 Data File : VN069174.D
 Acq On : 26 Oct 2021 00:51
 Operator : JC/MD
 Sample : M4313-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1-3-TOTES

Quant Time: Oct 26 01:38:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	365541	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	646708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	650159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	269143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	234627	48.683	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.360%
35) Dibromofluoromethane	8.027	113	188205	50.020	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.040%
50) Toluene-d8	10.446	98	810281	53.822	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.734	95	314273	55.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.660%
Target Compounds						
16) Acetone	4.299	43	18457	10.262	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102621\
Data File : VN069174.D
Acq On : 26 Oct 2021 00:51
Operator : JC/MD
Sample : M4313-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1-3-TOTES

Quant Time: Oct 26 01:38:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

