



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102621\
 Data File : VN069152.D
 Acq On : 25 Oct 2021 14:48
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Title : SW846 8260
 Last Update : Wed Oct 13 05:44:40 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	13459	PASS
75	95	30	60	54.6	36029	PASS
95	95	100	100	100.0	65957	PASS
96	95	5	9	7.2	4763	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.5	49800	PASS
175	174	5	9	7.5	3737	PASS
176	174	95	101	96.2	47883	PASS
177	176	5	9	6.1	2934	PASS