

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102622\
 Data File : VN075024.D
 Acq On : 26 Oct 2022 14:58
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN102622

Quant Time: Oct 27 09:55:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.650	0.580	10.8	91	0.00
3 P	Chloromethane	0.691	0.654	5.4	98	0.00
4 C	Vinyl Chloride	0.986	0.971	1.5#	99	0.00
5 T	Bromomethane	0.808	0.829	-2.6	105	0.00
6 T	Chloroethane	0.707	0.695	1.7	102	0.01
7 T	Trichlorofluoromethane	1.007	0.953	5.4	96	0.00
8 T	Diethyl Ether	0.367	0.356	3.0	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.556	0.532	4.3	98	0.00
10 T	Methyl Iodide	0.828	0.790	4.6	100	0.00
11 T	Tert butyl alcohol	0.157	0.148	5.7	103	0.00
12 CM	1,1-Dichloroethene	0.530	0.498	6.0#	96	0.00
13 T	Acrolein	0.129	0.123	4.7	100	0.00
14 T	Allyl chloride	0.774	0.773	0.1	98	0.01
15 T	Acrylonitrile	0.332	0.331	0.3	102	0.00
16 T	Acetone	0.285	0.322	-13.0	121	0.00
17 T	Carbon Disulfide	1.333	1.264	5.2	95	0.00
18 T	Methyl Acetate	0.932	0.915	1.8	101	0.00
19 T	Methyl tert-butyl Ether	1.970	1.913	2.9	98	0.00
20 T	Methylene Chloride	0.636	0.600	5.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.580	0.546	5.9	97	0.00
22 T	Diisopropyl ether	1.708	1.678	1.8	98	0.00
23 T	Vinyl Acetate	1.427	1.451	-1.7	100	0.00
24 P	1,1-Dichloroethane	1.080	1.059	1.9	100	0.00
25 T	2-Butanone	0.449	0.464	-3.3	107	0.00
26 T	2,2-Dichloropropane	0.994	1.011	-1.7	104	0.00
27 T	cis-1,2-Dichloroethene	0.693	0.688	0.7	101	0.01
28 T	Bromochloromethane	0.455	0.463	-1.8	101	0.00
29 T	Tetrahydrofuran	0.290	0.294	-1.4	102	0.00
30 C	Chloroform	1.173	1.160	1.1#	100	0.00
31 T	Cyclohexane	1.002	0.944	5.8	97	0.00
32 T	1,1,1-Trichloroethane	1.065	1.062	0.3	99	0.00
33 S	1,2-Dichloroethane-d4	0.674	0.694	-3.0	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.294	0.294	0.0	101	0.00
36 T	1,1-Dichloropropene	0.454	0.468	-3.1	101	0.00
37 T	Ethyl Acetate	0.498	0.489	1.8	102	0.00
38 T	Carbon Tetrachloride	0.500	0.512	-2.4	99	0.00
39 T	Methylcyclohexane	0.600	0.635	-5.8	103	0.00
40 TM	Benzene	1.391	1.415	-1.7	101	0.00
41 T	Methacrylonitrile	0.239	0.241	-0.8	99	0.00
42 TM	1,2-Dichloroethane	0.506	0.519	-2.6	102	0.00
43 T	Isopropyl Acetate	0.773	0.795	-2.8	101	0.00
44 TM	Trichloroethene	0.349	0.345	1.1	100	0.00
45 C	1,2-Dichloropropane	0.331	0.341	-3.0#	99	0.00
46 T	Dibromomethane	0.258	0.261	-1.2	102	0.00
47 T	Bromodichloromethane	0.500	0.514	-2.8	102	0.00
48 T	Methyl methacrylate	0.355	0.356	-0.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102622\
 Data File : VN075024.D
 Acq On : 26 Oct 2022 14:58
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN102622

Quant Time: Oct 27 09:55:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	109	0.00
50 S	Toluene-d8	1.139	1.227	-7.7	106	0.00
51 T	4-Methyl-2-Pentanone	0.499	0.520	-4.2	104	0.00
52 CM	Toluene	0.930	0.988	-6.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.527	0.576	-9.3	102	0.00
54 T	cis-1,3-Dichloropropene	0.559	0.592	-5.9	101	0.00
55 T	1,1,2-Trichloroethane	0.369	0.368	0.3	97	0.00
56 T	Ethyl methacrylate	0.563	0.609	-8.2	102	0.00
57 T	1,3-Dichloropropane	0.614	0.639	-4.1	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.217	0.231	-6.5	98	0.00
59 T	2-Hexanone	0.379	0.412	-8.7	108	0.00
60 T	Dibromochloromethane	0.389	0.421	-8.2	101	0.00
61 T	1,2-Dibromoethane	0.378	0.403	-6.6	101	0.00
62 S	4-Bromofluorobenzene	0.400	0.467	-16.8	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.307	0.300	2.3	102	0.00
65 PM	Chlorobenzene	1.067	1.074	-0.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.396	0.0	100	0.00
67 C	Ethyl Benzene	1.920	2.014	-4.9#	103	0.00
68 T	m/p-Xylenes	0.770	0.821	-6.6	101	0.00
69 T	o-Xylene	0.762	0.805	-5.6	103	0.00
70 T	Styrene	1.215	1.332	-9.6	104	0.00
71 P	Bromoform	0.296	0.313	-5.7	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.014	3.911	2.6	102	0.00
74 T	N-amyl acetate	1.485	1.432	3.6	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.309	1.238	5.4	104	0.00
76 T	1,2,3-Trichloropropane	1.062	0.991	6.7	94	0.00
77 T	Bromobenzene	0.884	0.840	5.0	102	0.00
78 T	n-propylbenzene	4.507	4.596	-2.0	106	0.00
79 T	2-Chlorotoluene	2.754	2.729	0.9	104	0.00
80 T	1,3,5-Trimethylbenzene	3.371	3.480	-3.2	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.414	0.417	-0.7	105	0.00
82 T	4-Chlorotoluene	2.754	2.729	0.9	104	0.00
83 T	tert-Butylbenzene	2.996	2.969	0.9	103	0.00
84 T	1,2,4-Trimethylbenzene	3.411	3.562	-4.4	107	0.00
85 T	sec-Butylbenzene	4.191	4.415	-5.3	107	0.00
86 T	p-Isopropyltoluene	3.461	3.740	-8.1	108	0.00
87 T	1,3-Dichlorobenzene	1.731	1.761	-1.7	105	0.00
88 T	1,4-Dichlorobenzene	1.713	1.761	-2.8	107	0.00
89 T	n-Butylbenzene	2.787	3.049	-9.4	107	0.00
90 T	Hexachloroethane	0.635	0.654	-3.0	105	0.00
91 T	1,2-Dichlorobenzene	1.720	1.720	0.0	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.264	2.6	102	0.00
93 T	1,2,4-Trichlorobenzene	0.740	0.761	-2.8	102	0.00
94 T	Hexachlorobutadiene	0.361	0.359	0.6	107	0.00
95 T	Naphthalene	2.847	2.888	-1.4	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102622\
Data File : VN075024.D
Acq On : 26 Oct 2022 14:58
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN102622

Quant Time: Oct 27 09:55:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.736	0.741	-0.7	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6