

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
 Data File : VN079691.D
 Acq On : 26 Oct 2023 11:16
 Operator : JC\MD
 Sample : VN1026MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1026MBL01

Quant Time: Oct 27 01:18:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	223518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	431761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	421899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	131090	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	194377	53.634	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.260%
35) Dibromofluoromethane	8.171	113	175112	56.437	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.880%
50) Toluene-d8	10.571	98	518039	46.143	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.280%
62) 4-Bromofluorobenzene	12.853	95	174160	48.339	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
Data File : VN079691.D
Acq On : 26 Oct 2023 11:16
Operator : JC\MD
Sample : VN1026MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1026MBL01

Quant Time: Oct 27 01:18:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

