

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
 Data File : VN079700.D
 Acq On : 26 Oct 2023 17:24
 Operator : JC\MD
 Sample : 04980-02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SEWER-SLUDGE

Quant Time: Oct 27 01:23:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	267287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	503072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	495615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	200870	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	209111	48.251	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.500%
35) Dibromofluoromethane	8.165	113	184708	51.091	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	10.571	98	601839	46.008	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.853	95	236275	56.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
Data File : VN079700.D
Acq On : 26 Oct 2023 17:24
Operator : JC\MD
Sample : 04980-02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SEWER-SLUDGE

Quant Time: Oct 27 01:23:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

