

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
 Data File : VN079690.D
 Acq On : 26 Oct 2023 10:42
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 01:17:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.716	0.651	9.1	84	0.00
3 P	Chloromethane	0.894	0.760	15.0	84	0.00
4 C	Vinyl Chloride	0.930	0.874	6.0#	90	0.00
5 T	Bromomethane	0.557	0.615	-10.4	115	0.00
6 T	Chloroethane	0.701	0.571	18.5	86	0.00
7 T	Trichlorofluoromethane	1.098	1.060	3.5	96	0.00
8 T	Diethyl Ether	0.391	0.394	-0.8	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.642	0.621	3.3	100	0.00
10 T	Methyl Iodide	0.557	0.699	-25.5#	106	0.00
11 T	Tert butyl alcohol	0.122	0.099	18.9	80	0.00
12 CM	1,1-Dichloroethene	0.591	0.570	3.6#	94	-0.01
13 T	Acrolein	0.121	0.139	-14.9	108	0.00
14 T	Allyl chloride	0.953	0.883	7.3	91	0.00
15 T	Acrylonitrile	0.354	0.354	0.0	94	0.00
16 T	Acetone	0.322	0.351	-9.0	103	0.00
17 T	Carbon Disulfide	1.843	1.517	17.7	83	0.00
18 T	Methyl Acetate	1.574	1.471	6.5	94	0.00
19 T	Methyl tert-butyl Ether	1.919	2.005	-4.5	96	0.00
20 T	Methylene Chloride	0.816	0.704	13.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.690	0.628	9.0	93	0.00
22 T	Diisopropyl ether	2.138	2.250	-5.2	98	0.00
23 T	Vinyl Acetate	1.202	1.240	-3.2	90	0.00
24 P	1,1-Dichloroethane	1.308	1.309	-0.1	99	0.00
25 T	2-Butanone	0.477	0.486	-1.9	93	0.00
26 T	2,2-Dichloropropane	1.000	1.061	-6.1	100	0.00
27 T	cis-1,2-Dichloroethene	0.750	0.765	-2.0	99	0.00
28 T	Bromochloromethane	0.622	0.626	-0.6	93	0.00
29 T	Tetrahydrofuran	0.309	0.311	-0.6	91	0.00
30 C	Chloroform	1.322	1.340	-1.4#	101	0.00
31 T	Cyclohexane	1.150	1.028	10.6	92	0.00
32 T	1,1,1-Trichloroethane	1.112	1.138	-2.3	100	0.00
33 S	1,2-Dichloroethane-d4	0.811	0.797	1.7	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.359	0.383	-6.7	101	0.00
36 T	1,1-Dichloropropene	0.539	0.588	-9.1	97	0.00
37 T	Ethyl Acetate	0.537	0.581	-8.2	93	0.00
38 T	Carbon Tetrachloride	0.540	0.578	-7.0	97	0.00
39 T	Methylcyclohexane	0.518	0.550	-6.2	90	0.00
40 TM	Benzene	1.602	1.742	-8.7	99	0.00
41 T	Methacrylonitrile	0.282	0.317	-12.4	94	0.00
42 TM	1,2-Dichloroethane	0.582	0.624	-7.2	99	0.00
43 T	Isopropyl Acetate	0.940	0.956	-1.7	93	0.00
44 TM	Trichloroethene	0.377	0.400	-6.1	98	0.00
45 C	1,2-Dichloropropane	0.422	0.467	-10.7#	102	0.00
46 T	Dibromomethane	0.273	0.305	-11.7	99	0.00
47 T	Bromodichloromethane	0.562	0.631	-12.3	100	0.00
48 T	Methyl methacrylate	0.403	0.443	-9.9	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
 Data File : VN079690.D
 Acq On : 26 Oct 2023 10:42
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 01:17:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	80	0.00
50 S	Toluene-d8	1.300	1.386	-6.6	95	0.00
51 T	4-Methyl-2-Pentanone	0.531	0.602	-13.4	95	0.00
52 CM	Toluene	0.943	1.073	-13.8#	98	0.00
53 T	t-1,3-Dichloropropene	0.567	0.658	-16.0	99	0.00
54 T	cis-1,3-Dichloropropene	0.620	0.712	-14.8	98	0.00
55 T	1,1,2-Trichloroethane	0.378	0.428	-13.2	102	0.00
56 T	Ethyl methacrylate	0.520	0.631	-21.3	95	0.00
57 T	1,3-Dichloropropane	0.654	0.746	-14.1	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.252	-26.6#	101	0.00
59 T	2-Hexanone	0.383	0.442	-15.4	94	0.00
60 T	Dibromochloromethane	0.380	0.454	-19.5	103	0.00
61 T	1,2-Dibromoethane	0.376	0.421	-12.0	100	0.00
62 S	4-Bromofluorobenzene	0.417	0.481	-15.3	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.344	0.358	-4.1	100	0.00
65 PM	Chlorobenzene	1.114	1.178	-5.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.408	0.456	-11.8	104	0.00
67 C	Ethyl Benzene	1.935	2.191	-13.2#	98	0.00
68 T	m/p-Xylenes	0.703	0.833	-18.5	101	0.00
69 T	o-Xylene	0.692	0.785	-13.4	98	0.00
70 T	Styrene	1.092	1.349	-23.5	100	0.00
71 P	Bromoform	0.298	0.331	-11.1	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	4.652	4.535	2.5	100	0.00
74 T	N-amyl acetate	1.907	1.843	3.4	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.814	1.527	15.8	101	0.00
76 T	1,2,3-Trichloropropane	1.442	1.361	5.6	98	0.00
77 T	Bromobenzene	1.141	1.038	9.0	100	0.00
78 T	n-propylbenzene	5.263	5.425	-3.1	101	0.00
79 T	2-Chlorotoluene	3.368	3.241	3.8	103	0.00
80 T	1,3,5-Trimethylbenzene	3.668	3.839	-4.7	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.497	0.471	5.2	97	0.00
82 T	4-Chlorotoluene	3.205	3.155	1.6	100	0.00
83 T	tert-Butylbenzene	3.005	3.047	-1.4	99	0.00
84 T	1,2,4-Trimethylbenzene	3.501	3.984	-13.8	108	0.00
85 T	sec-Butylbenzene	4.192	4.300	-2.6	99	0.00
86 T	p-Isopropyltoluene	3.216	3.474	-8.0	100	0.00
87 T	1,3-Dichlorobenzene	1.809	1.878	-3.8	102	0.00
88 T	1,4-Dichlorobenzene	1.868	1.824	2.4	101	0.00
89 T	n-Butylbenzene	2.568	2.924	-13.9	100	0.00
90 T	Hexachloroethane	0.756	0.693	8.3	103	0.00
91 T	1,2-Dichlorobenzene	1.765	1.816	-2.9	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.247	8.2	94	0.00
93 T	1,2,4-Trichlorobenzene	0.547	0.614	-12.2	91	0.00
94 T	Hexachlorobutadiene	0.426	0.436	-2.3	107	0.00
95 T	Naphthalene	1.669	1.597	4.3	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102623\
Data File : VN079690.D
Acq On : 26 Oct 2023 10:42
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 01:17:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.652	0.635	2.6	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6