

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102720\
 Data File : VN064339.D
 Acq On : 27 Oct 2020 20:25
 Operator : JC/MD
 Sample : L4511-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OB-9

Quant Time: Oct 28 05:05:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 12:23:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	378789	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	383460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166463	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	179125	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
35) Dibromofluoromethane	7.55	113	129287	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
50) Toluene-d8	10.06	98	484778	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.38	95	210056	57.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102720\
Data File : VN064339.D
Acq On : 27 Oct 2020 20:25
Operator : JC/MD
Sample : L4511-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OB-9

Quant Time: Oct 28 05:05:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 12:23:17 2020
Response via : Initial Calibration

