

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069182.D
 Acq On : 26 Oct 2021 13:49
 Operator : JC/MD
 Sample : VN1027WBSD02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 27 06:54:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	245996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	424820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	400104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	175624	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	175779	54.196	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.400%	
35) Dibromofluoromethane	8.024	113	132458	53.591	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.180%	
50) Toluene-d8	10.443	98	520390	52.621	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.240%	
62) 4-Bromofluorobenzene	12.734	95	200388	53.706	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.420%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	2.075	85	49543	17.015	ug/l 98
3) Chloromethane	2.303	50	49956	16.615	ug/l 97
4) Vinyl Chloride	2.445	62	59906	19.616	ug/l 98
5) Bromomethane	2.864	94	40302	23.991	ug/l 96
6) Chloroethane	3.025	64	40387	22.440	ug/l 99
7) Trichlorofluoromethane	3.382	101	86178	20.152	ug/l 95
8) Diethyl Ether	3.832	74	32141	20.149	ug/l 97
9) 1,1,2-Trichlorotrifluo...	4.218	101	47697	20.743	ug/l 99
10) Methyl Iodide	4.427	142	48853	14.638	ug/l 100
11) Tert butyl alcohol	5.366	59	63916	107.848	ug/l # 88
12) 1,1-Dichloroethene	4.191	96	43688	18.380	ug/l 96
13) Acrolein	4.047	56	10949	41.763	ug/l 99
14) Allyl chloride	4.851	41	78836	18.502	ug/l 96
15) Acrylonitrile	5.554	53	153084	115.277	ug/l 99
16) Acetone	4.288	43	152397	125.910	ug/l 99
17) Carbon Disulfide	4.537	76	101958	15.805	ug/l 100
18) Methyl Acetate	4.857	43	105789	23.043	ug/l 98
19) Methyl tert-butyl Ether	5.618	73	183843	20.772	ug/l 100
20) Methylene Chloride	5.101	84	55368	19.875	ug/l 98
21) trans-1,2-Dichloroethene	5.608	96	47513	18.502	ug/l 95
22) Diisopropyl ether	6.498	45	170292	19.218	ug/l 99
23) Vinyl Acetate	6.436	43	668790	99.835	ug/l 98
24) 1,1-Dichloroethane	6.391	63	93992	19.523	ug/l 100
25) 2-Butanone	7.335	43	216611	118.349	ug/l 97
26) 2,2-Dichloropropane	7.329	77	85993	18.742	ug/l 78
27) cis-1,2-Dichloroethene	7.324	96	61628	19.853	ug/l 98
28) Bromochloromethane	7.657	49	40683	18.657	ug/l 96
29) Tetrahydrofuran	7.686	42	137608	115.910	ug/l 97
30) Chloroform	7.820	83	103260	20.351	ug/l 95
31) Cyclohexane	8.099	56	88146	17.479	ug/l 94
32) 1,1,1-Trichloroethane	8.019	97	89650	18.954	ug/l 99
36) 1,1-Dichloropropene	8.225	75	73635	20.293	ug/l 99
37) Ethyl Acetate	7.412	43	85088	22.704	ug/l 99
38) Carbon Tetrachloride	8.209	117	74884	18.791	ug/l 98
39) Methylcyclohexane	9.462	83	89885	19.246	ug/l 97
40) Benzene	8.464	78	221349	20.033	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069182.D
 Acq On : 26 Oct 2021 13:49
 Operator : JC/MD
 Sample : VN1027WBSD02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 27 06:54:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.638	41	40489	21.433	ug/l	94
42) 1,2-Dichloroethane	8.531	62	84927	21.705	ug/l	99
43) Isopropyl Acetate	8.560	43	142452	21.543	ug/l	99
44) Trichloroethene	9.217	130	56954	20.595	ug/l	100
45) 1,2-Dichloropropane	9.494	63	57221	19.957	ug/l	97
46) Dibromomethane	9.580	93	42102	21.512	ug/l	99
47) Bromodichloromethane	9.765	83	80581	19.868	ug/l	100
48) Methyl methacrylate	9.561	41	65078	20.946	ug/l	95
49) 1,4-Dioxane	9.566	88	25845	447.235	ug/l	96
51) 4-Methyl-2-Pentanone	10.330	43	438473	114.362	ug/l	99
52) Toluene	10.507	92	146591	20.621	ug/l	99
53) t-1,3-Dichloropropene	10.719	75	91452	19.906	ug/l	100
54) cis-1,3-Dichloropropene	10.191	75	94914	19.857	ug/l	98
55) 1,1,2-Trichloroethane	10.899	97	60577	21.613	ug/l	99
56) Ethyl methacrylate	10.762	69	98462	20.923	ug/l	98
57) 1,3-Dichloropropane	11.047	76	100191	21.679	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.044	63	120190	99.067	ug/l	99
59) 2-Hexanone	11.087	43	328341	119.416	ug/l	99
60) Dibromochloromethane	11.240	129	60491	19.121	ug/l	98
61) 1,2-Dibromoethane	11.347	107	63439	22.067	ug/l	99
64) Tetrachloroethene	10.980	164	50658	19.655	ug/l	99
65) Chlorobenzene	11.771	112	159648	21.086	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.843	131	58318	19.680	ug/l	99
67) Ethyl Benzene	11.846	91	291006	20.692	ug/l	100
68) m/p-Xylenes	11.956	106	220177	41.299	ug/l	100
69) o-Xylene	12.283	106	107730	20.188	ug/l	98
70) Styrene	12.296	104	180843	20.818	ug/l	99
71) Bromoform	12.460	173	42707	17.991	ug/l #	99
73) Isopropylbenzene	12.581	105	287368	19.658	ug/l	100
74) N-amyl acetate	12.390	43	109394	19.753	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.830	83	94960	22.479	ug/l	100
76) 1,2,3-Trichloropropane	12.884	75	116081	31.064	ug/l #	1
77) Bromobenzene	12.862	156	67114	20.763	ug/l	97
78) n-propylbenzene	12.921	91	332809	20.259	ug/l	100
79) 2-Chlorotoluene	13.010	91	202378	19.987	ug/l	100
80) 1,3,5-Trimethylbenzene	13.061	105	240508	19.753	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.629	75	27263	17.455	ug/l	94
82) 4-Chlorotoluene	13.109	91	200788	20.521	ug/l	99
83) tert-Butylbenzene	13.326	119	210667	19.876	ug/l	99
84) 1,2,4-Trimethylbenzene	13.372	105	236824	20.187	ug/l	99
85) sec-Butylbenzene	13.503	105	298686	20.348	ug/l	99
86) p-Isopropyltoluene	13.619	119	241935	20.548	ug/l	100
87) 1,3-Dichlorobenzene	13.619	146	118827	21.252	ug/l	99
88) 1,4-Dichlorobenzene	13.696	146	118801	21.408	ug/l	98
89) n-Butylbenzene	13.946	91	205247	20.940	ug/l	99
90) Hexachloroethane	14.214	117	41371	16.727	ug/l	98
91) 1,2-Dichlorobenzene	13.991	146	112349	20.685	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.606	75	18106	22.230	ug/l	97
93) 1,2,4-Trichlorobenzene	15.268	180	57633	21.124	ug/l	99
94) Hexachlorobutadiene	15.373	225	31649	20.464	ug/l	98
95) Naphthalene	15.509	128	166504	22.683	ug/l	99
96) 1,2,3-Trichlorobenzene	15.702	180	56199	21.531	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069182.D
Acq On : 26 Oct 2021 13:49
Operator : JC/MD
Sample : VN1027WBSD02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Oct 27 06:54:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 27 06:54:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.088 min Scan# 2274

Delta R.T. 0.002 min

Lab File: VN069182.D





Acq: 26 Oct 2021 13:49

Tgt Ion:168 Resp: 245996

Ion Ratio Lower Upper

168 100

99 61.4 48.9 73.3



#2

Dichlorodifluoromethane

Concen: 17.015 ug/l

RT: 2.075 min Scan# 32

Delta R.T. 0.002 min

Lab File: VN069182.D




Acq: 26 Oct 2021 13:49


Tgt Ion: 85 Resp: 49543

Ion Ratio Lower Upper

85 100

87 31.4 16.3 48.8





#3

Chloromethane

Concen: 16.615 ug/l

RT: 2.303 min Scan# 117

Delta R.T. 0.003 min

Lab File: VN069182.D





Acq: 26 Oct 2021 13:49

Tgt Ion: 50 Resp: 49956

Ion Ratio Lower Upper

50 100

52 31.5 26.3 39.5



#4

Vinyl Chloride

Concen: 19.616 ug/l

RT: 2.445 min Scan# 170

Delta R.T. 0.003 min

Lab File: VN069182.D




Acq: 26 Oct 2021 13:49

Tgt Ion: 62 Resp: 59906

Ion Ratio Lower Upper

62 100

64 30.8 25.7 38.5









#5
Bromomethane
Concen: 23.991 ug/l
RT: 2.864 min Scan# 326
Delta R.T. 0.027 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 94 Resp: 40302
Ion Ratio Lower Upper
94 100
96 89.1 74.2 111.4



#6
Chloroethane
Concen: 22.440 ug/l
RT: 3.025 min Scan# 386
Delta R.T. 0.016 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 64 Resp: 40387
Ion Ratio Lower Upper
64 100
66 32.4 25.6 38.4









#7
 Trichlorofluoromethane
 Concen: 20.152 ug/l
 RT: 3.382 min Scan# 519
 Delta R.T. 0.011 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:101 Resp: 86178
 Ion Ratio Lower Upper
 101 100
 103 60.6 51.7 77.5



#8
 Diethyl Ether
 Concen: 20.149 ug/l
 RT: 3.832 min Scan# 687
 Delta R.T. 0.005 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49

Tgt Ion: 74 Resp: 32141
 Ion Ratio Lower Upper
 74 100
 45 91.1 47.0 141.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.743 ug/l
RT: 4.218 min Scan# 831
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 47697
Ion Ratio Lower Upper
101 100
85 44.6 37.2 55.8
151 76.4 61.1 91.7



#10
Methyl Iodide
Concen: 14.638 ug/l
RT: 4.427 min Scan# 909
Delta R.T. 0.005 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:142 Resp: 48853
Ion Ratio Lower Upper
142 100
127 43.0 34.4 51.6
141 13.9 11.3 16.9









#11

Tert butyl alcohol
Concen: 107.848 ug/l
RT: 5.366 min Scan# 1259
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Instrument :
MSVOA_N
ClientSampled :




Tgt Ion: 59 Resp: 63916
Ion Ratio Lower Upper
59 100
57 6.1 8.3 12.5#




#12

1,1-Dichloroethene
Concen: 18.380 ug/l
RT: 4.191 min Scan# 821
Delta R.T. 0.005 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 96 Resp: 43688
Ion Ratio Lower Upper
96 100
61 154.9 129.8 194.6
98 64.4 51.4 77.2









#13
Acrolein
Concen: 41.763 ug/l
RT: 4.047 min Scan# 767
Delta R.T. 0.005 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 10949
Ion Ratio Lower Upper
56 100
55 73.6 58.1 87.1



#14
Allyl chloride
Concen: 18.502 ug/l
RT: 4.851 min Scan# 1067
Delta R.T. 0.008 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 41 Resp: 78836
Ion Ratio Lower Upper
41 100
39 72.6 55.6 83.4
76 34.9 26.4 39.6









#15
Acrylonitrile
Concen: 115.277 ug/l
RT: 5.554 min Scan# 1329
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 53 Resp: 153084
Ion Ratio Lower Upper
53 100
52 83.7 65.7 98.5
51 36.0 28.9 43.3



#16
Acetone
Concen: 125.910 ug/l
RT: 4.288 min Scan# 857
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 43 Resp: 152397
Ion Ratio Lower Upper
43 100
58 31.3 24.5 36.7









#17
Carbon Disulfide
Concen: 15.805 ug/l
RT: 4.537 min Scan# 950
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 101958
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 23.043 ug/l
RT: 4.857 min Scan# 1069
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 43 Resp: 105789
Ion Ratio Lower Upper
43 100
74 24.6 18.9 28.3









#19

Methyl tert-butyl Ether
 Concen: 20.772 ug/l
 RT: 5.618 min Scan# 1353
 Delta R.T. 0.003 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49

Instrument :
 MSVOA_N
 ClientSampled :




Tgt Ion: 73 Resp: 183843
 Ion Ratio Lower Upper
 73 100
 57 21.4 17.3 25.9




#20

Methylene Chloride
 Concen: 19.875 ug/l
 RT: 5.101 min Scan# 1160
 Delta R.T. 0.005 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49

Tgt Ion: 84 Resp: 55368
 Ion Ratio Lower Upper
 84 100
 49 119.8 97.6 146.4
 51 37.3 30.3 45.5
 86 66.8 51.8 77.6








#21
trans-1,2-Dichloroethene
Concen: 18.502 ug/l
RT: 5.608 min Scan# 1349
Delta R.T. 0.008 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 47513
Ion Ratio Lower Upper
96 100
61 132.6 112.4 168.6
98 66.8 52.0 78.0



#22
Diisopropyl ether
Concen: 19.218 ug/l
RT: 6.498 min Scan# 1681
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 45 Resp: 170292
Ion Ratio Lower Upper
45 100
43 60.6 48.0 72.0
87 30.5 23.4 35.0
59 12.2 9.5 14.3





#23

Vinyl Acetate

Concen: 99.835 ug/l

RT: 6.436 min Scan# 1658

Delta R.T. 0.003 min

Lab File: VN069182.D





Acq: 26 Oct 2021 13:49

Tgt Ion: 43 Resp: 668790

Ion Ratio Lower Upper

43 100

86 11.1 8.3 12.5



#24

1,1-Dichloroethane

Concen: 19.523 ug/l

RT: 6.391 min Scan# 1641

Delta R.T. 0.003 min

Lab File: VN069182.D

Acq: 26 Oct 2021 13:49




Tgt Ion: 63 Resp: 93992


Ion Ratio Lower Upper

63 100

98 7.0 3.5 10.4

100 4.4 2.1 6.5





#25

2-Butanone

Concen: 118.349 ug/l

RT: 7.335 min Scan# 1993

Delta R.T. 0.003 min

Lab File: VN069182.D





Acq: 26 Oct 2021 13:49

Tgt Ion: 43 Resp: 216611

Ion Ratio Lower Upper

43 100

72 28.1 21.2 31.8



#26

2,2-Dichloropropane

Concen: 18.742 ug/l

RT: 7.329 min Scan# 1991

Delta R.T. 0.003 min

Lab File: VN069182.D




Acq: 26 Oct 2021 13:49

Tgt Ion: 77 Resp: 85993

Ion Ratio Lower Upper

77 100

97 12.0 11.2 33.6









#27

cis-1,2-Dichloroethene
Concen: 19.853 ug/l
RT: 7.324 min Scan# 1989
Delta R.T. -0.005 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Instrument :
MSVOA_N
ClientSampleId :




Tgt Ion: 96 Resp: 61628
Ion Ratio Lower Upper
96 100
61 141.3 0.0 290.6
98 63.1 0.0 127.2




#28

Bromochloromethane
Concen: 18.657 ug/l
RT: 7.657 min Scan# 2113
Delta R.T. -0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 49 Resp: 40683
Ion Ratio Lower Upper
49 100
129 1.7 0.0 4.4
130 79.4 60.8 91.2





#29

Tetrahydrofuran

Concen: 115.910 ug/l

RT: 7.686 min Scan# 2124





Delta R.T. 0.003 min

Lab File: VN069182.D

Acq: 26 Oct 2021 13:49

Tgt Ion: 42 Resp: 137608

Ion	Ratio	Lower	Upper
42	100		
72	45.8	35.4	53.2
71	44.6	33.8	50.6



#30

Chloroform

Concen: 20.351 ug/l

RT: 7.820 min Scan# 2174




Delta R.T. 0.003 min


Lab File: VN069182.D

Acq: 26 Oct 2021 13:49

Tgt Ion: 83 Resp: 103260

Ion	Ratio	Lower	Upper
83	100		
85	68.1	51.1	76.7









#31
Cyclohexane
Concen: 17.479 ug/l
RT: 8.099 min Scan# 2278
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 88146
Ion Ratio Lower Upper
56 100
69 35.0 26.9 40.3
84 90.6 67.6 101.4



#32
1,1,1-Trichloroethane
Concen: 18.954 ug/l
RT: 8.019 min Scan# 2248
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 97 Resp: 89650
Ion Ratio Lower Upper
97 100
99 62.7 50.8 76.2
61 44.0 34.6 51.8









#33
 1,2-Dichloroethane-d4
 Concen: 54.196 ug/l
 RT: 8.440 min Scan# 2405
 Delta R.T. 0.000 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 65 Resp: 175779
 Ion Ratio Lower Upper
 65 100
 67 51.8 0.0 103.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.968 min Scan# 2602
 Delta R.T. 0.000 min
 Lab File: VN069182.D
 Acq: 26 Oct 2021 13:49

Tgt Ion: 114 Resp: 424820
 Ion Ratio Lower Upper
 114 100
 63 20.9 0.0 42.6
 88 16.1 0.0 32.0









#35
Dibromofluoromethane
Concen: 53.591 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:113 Resp: 132458
Ion Ratio Lower Upper
113 100
111 101.6 83.2 124.8
192 18.3 15.0 22.4



#36
1,1-Dichloropropene
Concen: 20.293 ug/l
RT: 8.225 min Scan# 2325
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 75 Resp: 73635
Ion Ratio Lower Upper
75 100
110 35.4 17.9 53.8
77 31.0 24.7 37.1









#37
Ethyl Acetate
Concen: 22.704 ug/l
RT: 7.412 min Scan# 2022
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 85088
Ion Ratio Lower Upper
43 100
61 14.5 11.2 16.8
70 10.9 8.4 12.6



#38
Carbon Tetrachloride
Concen: 18.791 ug/l
RT: 8.209 min Scan# 2319
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 117 Resp: 74884
Ion Ratio Lower Upper
117 100
119 93.3 76.9 115.3
121 31.4 24.6 37.0









#39
Methylcyclohexane
Concen: 19.246 ug/l
RT: 9.462 min Scan# 2786
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 89885
Ion Ratio Lower Upper
83 100
55 72.9 61.4 92.0
98 45.8 36.4 54.6



#40
Benzene
Concen: 20.033 ug/l
RT: 8.464 min Scan# 2414
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 78 Resp: 221349
Ion Ratio Lower Upper
78 100
77 23.4 19.1 28.7









#41
Methacrylonitrile
Concen: 21.433 ug/l
RT: 7.638 min Scan# 2106
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion: 41 Resp: 40489




Ion	Ratio	Lower	Upper
41	100		
39	57.6	44.9	67.3
67	85.7	62.6	93.8
52	37.0	26.2	39.4




#42
1,2-Dichloroethane
Concen: 21.705 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 62 Resp: 84927

Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	17.2









#43
Isopropyl Acetate
Concen: 21.543 ug/l
RT: 8.560 min Scan# 2450
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 142452
Ion Ratio Lower Upper
43 100
61 25.8 20.4 30.6
87 13.9 10.7 16.1



#44
Trichloroethene
Concen: 20.595 ug/l
RT: 9.217 min Scan# 2695
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:130 Resp: 56954
Ion Ratio Lower Upper
130 100
95 104.1 0.0 208.0









#45
1,2-Dichloropropane
Concen: 19.957 ug/l
RT: 9.494 min Scan# 2798
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 57221
Ion Ratio Lower Upper
63 100
65 30.5 25.6 38.4



#46
Dibromomethane
Concen: 21.512 ug/l
RT: 9.580 min Scan# 2830
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 93 Resp: 42102
Ion Ratio Lower Upper
93 100
95 83.7 66.5 99.7
174 90.9 73.8 110.8









#47
Bromodichloromethane
Concen: 19.868 ug/l
RT: 9.765 min Scan# 2899
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 80581
Ion Ratio Lower Upper
83 100
85 63.5 50.7 76.1
127 8.4 6.5 9.7



#48
Methyl methacrylate
Concen: 20.946 ug/l
RT: 9.561 min Scan# 2823
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 41 Resp: 65078
Ion Ratio Lower Upper
41 100
69 89.3 66.4 99.6
39 61.3 50.2 75.2









#49
1,4-Dioxane
Concen: 447.235 ug/l
RT: 9.566 min Scan# 2825
Delta R.T. -0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 88 Resp: 25845
Ion Ratio Lower Upper
88 100
43 31.6 24.9 37.3
58 66.9 56.9 85.3



#50
Toluene-d8
Concen: 52.621 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 98 Resp: 520390
Ion Ratio Lower Upper
98 100
100 63.9 51.5 77.3









#51
4-Methyl-2-Pentanone
Concen: 114.362 ug/l
RT: 10.330 min Scan# 3110
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 438473
Ion Ratio Lower Upper
43 100
58 39.5 31.0 46.4



#52
Toluene
Concen: 20.621 ug/l
RT: 10.507 min Scan# 3176
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 92 Resp: 146591
Ion Ratio Lower Upper
92 100
91 174.3 138.5 207.7









#53
t-1,3-Dichloropropene
Concen: 19.906 ug/l
RT: 10.719 min Scan# 3255
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 91452
Ion Ratio Lower Upper
75 100
77 32.0 25.4 38.2



#54
cis-1,3-Dichloropropene
Concen: 19.857 ug/l
RT: 10.191 min Scan# 3058
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 75 Resp: 94914
Ion Ratio Lower Upper
75 100
77 30.8 25.6 38.4
39 51.7 42.9 64.3









#55
1,1,2-Trichloroethane
Concen: 21.613 ug/l
RT: 10.899 min Scan# 3322
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:	97	Resp:	60577
Ion	Ratio	Lower	Upper
97	100		
83	88.7	70.5	105.7
85	55.6	45.3	67.9
99	62.5	50.7	76.1



#56
Ethyl methacrylate
Concen: 20.923 ug/l
RT: 10.762 min Scan# 3271
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:	69	Resp:	98462
Ion	Ratio	Lower	Upper
69	100		
41	64.4	53.3	79.9
39	41.3	33.3	49.9









#57
1,3-Dichloropropane
Concen: 21.679 ug/l
RT: 11.047 min Scan# 3377
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 100191
Ion Ratio Lower Upper
76 100
78 32.2 25.8 38.8



#58
2-Chloroethyl Vinyl ether
Concen: 99.067 ug/l
RT: 10.044 min Scan# 3003
Delta R.T. 0.001 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 63 Resp: 120190
Ion Ratio Lower Upper
63 100
106 28.3 22.2 33.2









#59
2-Hexanone
Concen: 119.416 ug/l
RT: 11.087 min Scan# 3392
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 328341
Ion Ratio Lower Upper
43 100
58 54.3 26.9 80.7



#60
Dibromochloromethane
Concen: 19.121 ug/l
RT: 11.240 min Scan# 3449
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 129 Resp: 60491
Ion Ratio Lower Upper
129 100
127 76.4 39.1 117.2









#61
1,2-Dibromoethane
Concen: 22.067 ug/l
RT: 11.347 min Scan# 3489
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:107 Resp: 63439
Ion Ratio Lower Upper
107 100
109 94.5 74.7 112.1



#62
4-Bromofluorobenzene
Concen: 53.706 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 95 Resp: 200388
Ion Ratio Lower Upper
95 100
174 73.2 0.0 144.8
176 71.6 0.0 140.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:117 Resp: 400104
Ion Ratio Lower Upper
117 100
82 55.7 46.4 69.6
119 31.4 26.0 39.0



#64
Tetrachloroethene
Concen: 19.655 ug/l
RT: 10.980 min Scan# 3352
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:164 Resp: 50658
Ion Ratio Lower Upper
164 100
166 126.6 99.8 149.8
129 100.8 79.2 118.8
131 96.3 76.3 114.5








#65
Chlorobenzene
Concen: 21.086 ug/l
RT: 11.771 min Scan# 3647
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:112 Resp: 159648
Ion Ratio Lower Upper
112 100
114 31.3 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 19.680 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:131 Resp: 58318
Ion Ratio Lower Upper
131 100
133 95.8 47.3 142.0
119 64.8 33.0 99.0









#67
Ethyl Benzene
Concen: 20.692 ug/l
RT: 11.846 min Scan# 3675
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 291006
Ion Ratio Lower Upper
91 100
106 31.0 24.8 37.2



#68
m/p-Xylenes
Concen: 41.299 ug/l
RT: 11.956 min Scan# 3716
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:106 Resp: 220177
Ion Ratio Lower Upper
106 100
91 208.2 166.8 250.2









#69
o-Xylene
Concen: 20.188 ug/l
RT: 12.283 min Scan# 3838
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 107730
Ion Ratio Lower Upper
106 100
91 222.0 109.6 328.6



#70
Styrene
Concen: 20.818 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:104 Resp: 180843
Ion Ratio Lower Upper
104 100
78 52.1 42.2 63.4
103 55.2 44.4 66.6









#71
Bromoform
Concen: 17.991 ug/l
RT: 12.460 min Scan# 3904
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:173 Resp: 42707
Ion Ratio Lower Upper
173 100
175 48.7 23.8 71.5
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:152 Resp: 175624
Ion Ratio Lower Upper
152 100
115 64.0 31.4 94.3
150 165.4 0.0 351.0









#73
Isopropylbenzene
Concen: 19.658 ug/l
RT: 12.581 min Scan# 3949
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 287368
Ion Ratio Lower Upper
105 100
120 26.0 13.1 39.3



#74
N-ethyl acetate
Concen: 19.753 ug/l
RT: 12.390 min Scan# 3878
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 43 Resp: 109394
Ion Ratio Lower Upper
43 100
70 41.7 32.2 48.2
55 23.5 18.0 27.0
61 24.3 19.0 28.4









#75

1,1,2,2-Tetrachloroethane
Concen: 22.479 ug/l
RT: 12.830 min Scan# 4042
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Instrument :
MSVOA_N
ClientSampled :




Tgt Ion: 83 Resp: 94960
Ion Ratio Lower Upper
83 100
131 10.4 5.1 15.2
85 64.3 32.3 96.8




#76

1,2,3-Trichloropropane
Concen: 31.064 ug/l
RT: 12.884 min Scan# 4062
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 75 Resp: 116081
Ion Ratio Lower Upper
75 100
77 0.0 96.2 288.4#









#77
Bromobenzene
Concen: 20.763 ug/l
RT: 12.862 min Scan# 4054
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:156 Resp: 67114
Ion Ratio Lower Upper
156 100
77 217.7 111.8 335.3
158 96.4 48.4 145.0



#78
n-propylbenzene
Concen: 20.259 ug/l
RT: 12.921 min Scan# 4076
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 91 Resp: 332809
Ion Ratio Lower Upper
91 100
120 22.7 11.4 34.1









#79
2-Chlorotoluene
Concen: 19.987 ug/l
RT: 13.010 min Scan# 4109
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 202378
Ion Ratio Lower Upper
91 100
126 33.0 16.4 49.4



#80
1,3,5-Trimethylbenzene
Concen: 19.753 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 105 Resp: 240508
Ion Ratio Lower Upper
105 100
120 47.1 23.4 70.2









#81
trans-1,4-Dichloro-2-butene
Concen: 17.455 ug/l
RT: 12.629 min Scan# 3967
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 27263
Ion Ratio Lower Upper
75 100
53 102.9 75.4 113.2
89 46.0 36.9 55.3



#82
4-Chlorotoluene
Concen: 20.521 ug/l
RT: 13.109 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 91 Resp: 200788
Ion Ratio Lower Upper
91 100
126 32.8 16.3 48.8









#83
tert-Butylbenzene
Concen: 19.876 ug/l
RT: 13.326 min Scan# 4227
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 210667
Ion Ratio Lower Upper
119 100
91 67.8 34.6 103.9
134 25.2 12.6 37.6



#84
1,2,4-Trimethylbenzene
Concen: 20.187 ug/l
RT: 13.372 min Scan# 4244
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:105 Resp: 236824
Ion Ratio Lower Upper
105 100
120 44.4 21.9 65.5





#85

sec-Butylbenzene

Concen: 20.348 ug/l

RT: 13.503 min Scan# 4293

Delta R.T. 0.000 min

Lab File: VN069182.D

Acq: 26 Oct 2021 13:49

Tgt Ion:105 Resp: 298686

Ion Ratio Lower Upper





105 100

134 19.2 9.8 29.4

Instrument :

MSVOA_N

ClientSampled :



#86

p-Isopropyltoluene

Concen: 20.548 ug/l

RT: 13.619 min Scan# 4336

Delta R.T. 0.000 min

Lab File: VN069182.D

Acq: 26 Oct 2021 13:49




Tgt Ion:119 Resp: 241935


Ion Ratio Lower Upper

119 100

134 25.8 12.9 38.6

91 25.4 12.8 38.4









#87
1,3-Dichlorobenzene
Concen: 21.252 ug/l
RT: 13.619 min Scan# 4336
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:146 Resp: 118827
Ion Ratio Lower Upper
146 100
111 42.4 21.4 64.2
148 63.4 32.0 96.2



#88
1,4-Dichlorobenzene
Concen: 21.408 ug/l
RT: 13.696 min Scan# 4365
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:146 Resp: 118801
Ion Ratio Lower Upper
146 100
111 42.9 20.6 61.8
148 62.8 32.1 96.5









#89
n-Butylbenzene
Concen: 20.940 ug/l
RT: 13.946 min Scan# 4458
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 205247
Ion Ratio Lower Upper
91 100
92 51.4 26.0 78.0
134 25.2 12.5 37.5



#90
Hexachloroethane
Concen: 16.727 ug/l
RT: 14.214 min Scan# 4558
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 117 Resp: 41371
Ion Ratio Lower Upper
117 100
201 72.9 37.1 111.5









#91
1,2-Dichlorobenzene
Concen: 20.685 ug/l
RT: 13.991 min Scan# 4475
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:146 Resp: 112349
Ion Ratio Lower Upper
146 100
111 43.8 22.0 66.0
148 64.3 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.230 ug/l
RT: 14.606 min Scan# 4704
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion: 75 Resp: 18106
Ion Ratio Lower Upper
75 100
155 75.9 38.8 116.4
157 97.0 50.2 150.6









#93
1,2,4-Trichlorobenzene
Concen: 21.124 ug/l
RT: 15.268 min Scan# 4951
Delta R.T. 0.003 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:180 Resp: 57633
Ion Ratio Lower Upper
180 100
182 97.5 47.9 143.7
145 32.4 16.4 49.2



#94
Hexachlorobutadiene
Concen: 20.464 ug/l
RT: 15.373 min Scan# 4990
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:225 Resp: 31649
Ion Ratio Lower Upper
225 100
223 63.3 31.5 94.5
227 60.6 31.6 94.7









#95
Naphthalene
Concen: 22.683 ug/l
RT: 15.509 min Scan# 5041
Delta R.T. 0.002 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128 Resp: 166504
Ion Ratio Lower Upper
128 100
127 13.0 10.5 15.7
129 10.8 8.9 13.3



#96
1,2,3-Trichlorobenzene
Concen: 21.531 ug/l
RT: 15.702 min Scan# 5113
Delta R.T. 0.000 min
Lab File: VN069182.D
Acq: 26 Oct 2021 13:49

Tgt Ion:180 Resp: 56199
Ion Ratio Lower Upper
180 100
182 93.9 48.1 144.3
145 34.1 17.4 52.2

