

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069183.D
 Acq On : 26 Oct 2021 14:14
 Operator : JC/MD
 Sample : M4337-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DDC-5-PS

Quant Time: Oct 27 06:54:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	345102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	597484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	575868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	220627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225810	49.628	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.260%
35) Dibromofluoromethane	8.024	113	174453	50.185	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.360%
50) Toluene-d8	10.446	98	740416	53.233	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.460%
62) 4-Bromofluorobenzene	12.734	95	269070	51.274	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.540%
Target Compounds						
16) Acetone	4.309	43	3880	2.285	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069183.D
Acq On : 26 Oct 2021 14:14
Operator : JC/MD
Sample : M4337-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DDC-5-PS

Quant Time: Oct 27 06:54:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

