

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069187.D
 Acq On : 26 Oct 2021 15:54
 Operator : JC/MD
 Sample : M4337-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DDC-6-PD

Quant Time: Oct 27 06:56:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	337186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	587231	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	560296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	211221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222242	49.991	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.980%
35) Dibromofluoromethane	8.027	113	172868	50.597	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.200%
50) Toluene-d8	10.443	98	729190	53.341	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.680%
62) 4-Bromofluorobenzene	12.733	95	259533	50.320	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.640%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	4391	2.647	ug/l	96
30) Chloroform	7.817	83	2069	0.297	ug/l	85
64) Tetrachloroethene	10.979	164	9147	2.534	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069187.D
Acq On : 26 Oct 2021 15:54
Operator : JC/MD
Sample : M4337-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DDC-6-PD

Quant Time: Oct 27 06:56:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

