

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069188.D
 Acq On : 26 Oct 2021 16:20
 Operator : JC/MD
 Sample : M4337-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3S(OFFSITE)

Quant Time: Oct 27 06:56:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	339599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	596621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	569300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	213529	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	220843	49.323	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.640%
35) Dibromofluoromethane	8.027	113	174978	50.409	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.820%
50) Toluene-d8	10.443	98	738959	53.205	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.733	95	267416	51.033	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069188.D
Acq On : 26 Oct 2021 16:20
Operator : JC/MD
Sample : M4337-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3S(OFFSITE)

Quant Time: Oct 27 06:56:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

