

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069191.D
 Acq On : 26 Oct 2021 17:35
 Operator : JC/MD
 Sample : M4337-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2S(OFFSITE)

Manual Integrations
 APPROVED

MMDadoda
 10/27/2021 5:05:05 PM

Quant Time: Oct 27 06:57:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	339265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.970	114	589162	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	566969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	212656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	220126	49.211	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 98.420%			
35) Dibromofluoromethane	8.024	113	171735	50.101	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.200%			
50) Toluene-d8	10.443	98	730109	53.234	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.460%			
62) 4-Bromofluorobenzene	12.733	95	265694	51.346	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.700%			
Target Compounds						Qvalue
16) Acetone	4.304	43	3094m	1.854	ug/l	
27) cis-1,2-Dichloroethene	7.329	96	1998	0.467	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069191.D
Acq On : 26 Oct 2021 17:35
Operator : JC/MD
Sample : M4337-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2S(OFFSITE)

Manual Integrations
APPROVED

MMDadoda
10/27/2021 5:05:05 PM

Quant Time: Oct 27 06:57:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

