

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
 Data File : VN069204.D
 Acq On : 26 Oct 2021 23:53
 Operator : JC/MD
 Sample : M4337-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DDC-9-PS

Quant Time: Oct 27 09:25:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	327046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	559884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	542984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	196420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	213253	49.456	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.920%
35) Dibromofluoromethane	8.024	113	167449	51.405	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.800%
50) Toluene-d8	10.446	98	704560	54.057	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.120%
62) 4-Bromofluorobenzene	12.734	95	252061	51.259	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.520%
Target Compounds						
16) Acetone	4.307	43	5087	3.161	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102721\
Data File : VN069204.D
Acq On : 26 Oct 2021 23:53
Operator : JC/MD
Sample : M4337-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DDC-9-PS

Quant Time: Oct 27 09:25:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

